



ISO/IEC 17025:2017

מעבדות בדיקה

תעודת הסמכה מס' 507

מעבדות הוברמן בע"מ

כתובת אתר ייחוס: קיבוץ רמת השופט, 1923800

עד יום: 24.01.2025

בתוקף מיום: 05.12.2022

הארגון נבדק ונבחן על ידי הרשות הלאומית להסמכת מעבדות (להלן הרשות) ונמצא ראוי להסמכה בהתאם לנספח פירוט היקף ההסמכה המצורף לתעודה זו, המהווה חלק בלתי נפרד ממנה ומספרו זהה למספר התעודה. הסמכה מצביעה על כשירות מקצועית ותפעול מערכת ניהול איכות בעלת הכרה בינלאומית. הארגון המוסמך על ידי הרשות, עומד בתקנים/ בדרישות המפורטים מעלה. דרישות התקנים הם לכשירות מקצועית ולמערכות ניהול, שהינן הכרחיות למתן תוצאות אמינות. הסמכה זו ניתנה בהתאם לכללי ISO/IEC 17011:2017 לפיהם פועלת הרשות ובמסגרתם מקיימת פיקוח שוטף על הארגון לצורך בחינת תפקודו המתמשך בהתאם לדרישות ההסמכה. ההסמכה תקפה כל עוד הארגון עונה לאמות המידה שנקבעו על ידי הרשות. הרשות חתומה על הסכם הכרה רב צדדי (MLA) מול ארגון European Accreditation Cooperation (EA).

תעודה זו אינה מהווה אישור לפי סעיף 12 לחוק התקנים.

אתי פלר
מנכ"ל
הרשות הלאומית להסמכת מעבדות

תאריך הסמכה ראשון: 25.01.2021



הרשות הלאומית להסמכת מעבדות
Israel Laboratory Accreditation Authority

Testing Laboratories

ISO/IEC 17025:2017

Accreditation Certificate No. 507

Huberman Labs Ltd

Main site address: Kibutz Ramat Hashofet ,1923800, Israel

Valid from: 05.12.2022

Until: 24.01.2025

The organization was assessed by the Israel Laboratory Accreditation Authority (ISRAC) and found to be worthy of accreditation to the detailed schedule attached.

The schedule is an integral part of this certificate and is numbered with the above certificate number.

Accreditation demonstrates technical competence and operation of an internationally recognized quality management system.

The organization accredited by ISRAC complies with the standards/requirements mentioned above, meets the technical competence requirements and management system requirements that are necessary for it to consistently deliver technically competent results. This accreditation is granted in accordance with the requirements of ISO/IEC 17011:2017, and entails periodic surveillance and reassessment by ISRAC to ensure that the organization continues to comply with the accreditation requirements.

The accreditation is valid provided that the organization continues to meet the criteria as laid down by ISRAC. ISRAC is an EA-MLA (European Accreditation Cooperation Multi-Lateral Agreement) signatory.

This certificate does not constitute an approval in accordance with article 12 of the standard law.

Date of first accreditation: 25.01.2021

Etty Feller
General Manager
Israel Laboratory Accreditation Authority

Date of signature 05/12/2022

Page No. 2 of: 4



Name and Address:

| | |
|-------------------------|--|
| Organization name | Huberman Labs Ltd |
| Address | Kibutz Ramat Hashofet, 1923800, Israel |
| Phone | 972-04-6417087 |
| Fax | |
| E-mail (contact person) | marina@hubermanlabs.com |

Site: P or T or M , P-Permanent, T-Temporary, M-Mobile

A permanent (P) or temporary (T) place, or a stationary or mobile (M) facility, at or from which the organization performs activities forming part of its scope of accreditation, starting from sampling to final issuance of a report or certificate and / or quality system activities. A temporary (T) site is a site established under the responsibility of an accredited permanent site. All activities performed at a temporary site are the responsibility of the permanent site. An outdoors work is also considered to be a temporary site. Temporary site will be a site that involves work for special project and the activity will be defined in time (up to 2 years).

Type of Scopes: A- Fixed, C- Flexible scope in analytical tests : Type of matrix, analytes, experimental systems and/or analytical characteristics may be subject to changes, in accordance with the laboratory's approved and documented procedures. For details, please refer to the list of Accredited Tests, available from the laboratory upon request.



| Item | Scope Type | Site | Materials / Products Tested | Types of Test / Properties Measured | Standard / Method | Opinion and Interpretation | Remarks | | |
|---|------------|------|---|---|---|----------------------------|--------------------|------|-----------------------|
| Group of products: Fire & Rescue - Fire Rescue Systems | | | | | משפחת מוצרים: כבאות והצלה- מערכות הצלה | | | | |
| Chemical Testing, Classical Chemistry | | | | | בדיקות כימיות, כימיה קלאסית | | | | |
| 1 | A | P | Fire extinguishing media: Foam concentrates – low expansion foam concentrates | אמצעים לכיבוי אש: תרכיזי קצף בעל יחס התפשטות נמוך | Expansion rate | יחס התפשטות | SI 71568 Part 3 | ---- | |
| 2 | A | P | Foam Concentrate for fire fighting | תרכיזי קצף לכיבוי אש | Expansion rate | יחס התפשטות | EN 1568-3 Annex G | ---- | |
| 3 | A | P | Foam Concentrate for fire fighting | תרכיזי קצף לכיבוי אש | Specific Gravity | צפיפות סגולית | In house procedure | ---- | Based on: ISO 00649-2 |
| 4 | A | P | Fire extinguishing media: Foam concentrates | אמצעים לכיבוי אש: תרכיזי קצף | 25% drainage time | 25% זמן ניקוז | SI 71568 Part 3 | ---- | |
| 5 | A | P | Foam Concentrate for fire fighting | תרכיזי קצף לכיבוי אש | 25% drainage time | 25% זמן ניקוז | EN 1568-3 Annex G | ---- | |
| 6 | A | P | Foam Concentrate for fire fighting | תרכיזי קצף לכיבוי אש | pH | pH | In house procedure | ---- | Based on: ISO 10523 |