



ISO/IEC 17025:2017

מעבדות כיוול

תעודת הסמכה מס' 0593 סלומון ליון ואלשטיין (סלא) בע"מ

כתובת אתר ייחוס: התאנה 1, פארק תעשיות חבל מודיעין, ת.ד. 975 שוהם, 6085001

עד יום: 03.11.2028

בתוקף מיום: 11.10.2026

הארגון נבדק ונבחן על ידי הרשות הלאומית להסמכת מעבדות (להלן הרשות) ונמצא ראוי להסמכה בהתאם לנספח פירוט היקף ההסמכה המצורף לתעודה זו, המהווה חלק בלתי נפרד ממנה ומספרו זהה למספר התעודה. הסמכה מצביעה על כשירות מקצועית ותפעול מערכת ניהול איכות בעלת הכרה בינלאומית. הארגון המוסמך על ידי הרשות, עומד בתקנים/ בדרישות המפורטים מעלה. דרישות התקנים הם לכשירות מקצועית ולמערכות ניהול, שהינן הכרחיות למתן תוצאות אמינות. הסמכה זו ניתנה בהתאם לכללי ISO/IEC 17011:2017 לפיהם פועלת הרשות ובמסגרתם מקיימת פיקוח שוטף על הארגון לצורך בחינת תפקודו המתמשך בהתאם לדרישות ההסמכה. ההסמכה תקפה כל עוד הארגון עונה לאמות המידה שנקבעו על ידי הרשות. הרשות חתומה על הסכם הכרה רב צדדי (MLA) מול ארגון (EA) European Accreditation Cooperation.

תעודה זו אינה מהווה אישור לפי סעיף 12 לחוק התקנים.

אתי פלר
מנכ"ל

הרשות הלאומית להסמכת מעבדות

תאריך הסמכה ראשון: 04.11.2024



הרשות הלאומית להסמכת מעבדות
Israel Laboratory Accreditation Authority

Calibration Laboratories

ISO/IEC 17025:2017

Accreditation Certificate No.0593 Salomon Levin & Elstein (S.L.E) Ltd.

Main site address 1 Hate'ena St. Modiin industrial park P.O.B 875 Shoham 6085001

Valid from: 11.10.2026

Until: 03.11.2028

The organization was assessed by the Israel Laboratory Accreditation Authority (ISRAC) and found to be worthy of accreditation to the detailed schedule attached.

The schedule is an integral part of this certificate and is numbered with the above certificate number.

Accreditation demonstrates technical competence and operation of an internationally recognized quality management system.

The organization accredited by ISRAC complies with the standards/requirements mentioned above, meets the technical competence requirements and management system requirements that are necessary for it to consistently deliver technically competent results. This accreditation is granted in accordance with the requirements of ISO/IEC 17011:2017, and entails periodic surveillance and reassessment by ISRAC to ensure that the organization continues to comply with the accreditation requirements.

The accreditation is valid provided that the organization continues to meet the criteria as laid down by ISRAC. ISRAC is an EA-MLA (European Accreditation Cooperation Multi-Lateral Agreement) signatory.

This certificate does not constitute an approval in accordance with article 12 of the standard law.

Date of first accreditation: 04.11.2024

Etty Feller
General Manager
Israel Laboratory Accreditation Authority

Date of signature 11/10/2026

Page No. 2 of: 5



Name and Address:

Organization name	<u>Salomon Levin & Elstein (S.L.E) Ltd.</u>
Address	1 Hate'ena St. Modiin industrial park P.O.B 875 Shoham 6085001
Phone	03-6864211
Fax	03-6864227
E-mail (contact person)	Eyal.Turgeman@sle.co.il

Site: P or T or M , P-Permanent, T-Temporary, M-Mobile

A permanent (P) or temporary (T) place, or a stationary or mobile (M) facility, at or from which the organization performs activities forming part of its scope of accreditation, starting from sampling to final issuance of a report or certificate and / or quality system activities. A temporary (T) site is a site established under the responsibility of an accredited permanent site. All activities performed at a temporary site are the responsibility of the permanent site. An outdoors work is also considered to be a temporary site. Temporary site will be a site that involves work for special project and the activity will be defined in time (up to 2 years).

Type of Scopes: A- Fixed, C- Flexible scope in analytical tests : Type of matrix, analytes, experimental systems and/or analytical characteristics may be subject to changes, in accordance with the laboratory's approved and documented procedures. For details, please refer to the list of Accredited Tests, available from the laboratory upon request.



Item	Scope Type	Site	Measurand Instrument, Gauge	Range [Including margins] (Does not include margins)	Uncertainty of Measurement ¹	Reference Documents	Remarks
Calibration - Physical quantities - Humidity					כיוול - גדלים פיזיקליים – לחות		
1	A	P;T	Relative Humidity, Calibration of humidity instruments in ambient air	לחות יחסית, כיוול מדי לחות באוויר סביבתי [10 °C to 30 °C] [15 %rh to 75 %rh]	2.0 %rh	SOPW023118 Manufacturer's Specification Customer requirements	Reference temperature and humidity probe. Single point calibration in ambient conditions

Item	Scope Type	Site	Measurand Instrument, Gauge	Range [Including margins] (Does not include margins)	Uncertainty of Measurement ¹	Reference Documents	Remarks
Calibration - Physical quantities - Pressure					כיוול - גדלים פיזיקליים – לחץ		
2	A	P;T	Pressure, Pneumatic, differential	לחץ, לחץ יחסי פנאומטי, לחץ הפרשי [-0.25 kPa to 0.25 kPa]	0.5 Pa	SOPW023116 Based on:	Pressure calibrator. Calibration can be given in other units as required.
3	A	P;T	Pressure gauges and transducers	מדי לחץ ומתמרי לחץ [-2.0 kPa to 2.0 kPa]	1.0 Pa	EURAMET/cg-17 Manufacturer's Specification	

Item	Scope Type	Site	Measurand Instrument, Gauge	Range [Including margins] (Does not include margins)	Uncertainty of Measurement ¹	Reference Documents	Remarks
Calibration - Physical quantities - Mass					כיוול - גדלים פיזיקליים – מסה		
4	A	P;T	Mass, Non automatic weighing instruments	מסה, מאזניים לא אוטומטיים Up to 30 kg	$2 * \sqrt{\left(\frac{Res}{\sqrt{12}}\right)^2 + \left(\frac{MPE}{\sqrt{12}}\right)^2}$	SOPW023117 Based on: OIML-R76-1 EURAMET/cg-18	Res- Resolution of the balance. MPE- Maximum Permissible Error of reference weights (OIML-R- 111). Available standard weights are: F1 from 1 mg to 30 kg,



Item	Scope Type	Site	Measurand Instrument, Gauge		Range [Including margins] (Does not include margins)	Uncertainty of Measurement ¹	Reference Documents	Remarks
Calibration - Physical quantities - Temperature						כיוול - גדלים פיזיקליים – טמפרטורה		
5	A	P;T	Temperature,	טמפרטורה,	[-90 °C to -35 °C)	0.15 °C	SOPW023115	Calibration bath, PRT
6	A	P;T	Thermometers with sensor	מדי טמפרטורה עם רגש	[-35 °C to 50 °C]	0.07 °C	Based on: ITS-90 ASTM E2877 נוהל משרד הבריאות 126 Manufacturer's Specification	
7	A	P;T	Temperature, Calibration of temperature instruments in ambient air	טמפרטורה, כיוול מדי טמפרטורה באוויר סביבתי	[10 °C to 30 °C]	0.3 °C	SOPW023115 Manufacturer instructions Customer requirements	Reference temperature probe. Single point calibration in ambient conditions
8	A	P;T	Temperature, Temperature indicators and controllers in Temperature controlled chambers	טמפרטורה, מדי ובקרי טמפרטורה בתאים מבוקרים טמפרטורה	[-80 °C to 50 °C]	0.4 °C	SOPW023115 Based on: DKD-R-5-7 Manufacturer's Specification	Thermometer with PRT

¹⁾ The uncertainty covered by the CMC expressed as the standard measurement uncertainty multiplied by the coverage factor k such that the coverage probability corresponds to approximately 95 %.