



ISO/IEC 17025:2017

מעבדות כיוול

תעודת הסמכה מס' 0008

## אלביט מערכות ל"א וסיגינט - אלישרא בע"מ

כתובת אתר ייחוס: המרכבה 29, חולון, 5885118

עד יום: 23.11.2027

בתוקף מיום: 04.09.2025

הארגון נבדק ונבחן על ידי הרשות הלאומית להסמכת מעבדות (להלן הרשות) ונמצא ראוי להסמכה בהתאם לנספח פירוט היקף ההסמכה המצורף לתעודה זו, המהווה חלק בלתי נפרד ממנה ומספרו זהה למספר התעודה. הסמכה מצביעה על כשירות מקצועית ותפעול מערכת ניהול איכות בעלת הכרה בינלאומית. הארגון המוסמך על ידי הרשות, עומד בתקנים/ בדרישות המפורטים מעלה. דרישות התקנים הם לכשירות מקצועית ולמערכות ניהול, שהינן הכרחיות למתן תוצאות אמינות. הסמכה זו ניתנה בהתאם לכללי ISO/IEC 17011:2017 לפיהם פועלת הרשות ובמסגרתם מקיימת פיקוח שוטף על הארגון לצורך בחינת תפקודו המתמשך בהתאם לדרישות ההסמכה. ההסמכה תקפה כל עוד הארגון עונה לאמות המידה שנקבעו על ידי הרשות. הרשות חתומה על הסכם הכרה רב צדדי (MLA) מול ארגון (EA) European Accreditation Cooperation.

תעודה זו אינה מהווה אישור לפי סעיף 12 לחוק התקנים.

אתי פלר  
מנכ"ל  
הרשות הלאומית להסמכת מעבדות

תאריך הסמכה ראשון: 24.11.1997



הרשות הלאומית להסמכת מעבדות  
Israel Laboratory Accreditation Authority

**Calibration Laboratories**

**ISO/IEC 17025:2017**

**Accreditation Certificate No.0008**

**Elbit Systems EW and SIGINT Elisra Ltd.**

**Main site address:** 29 Hamerkava , Holon, 5885118

**Valid from: 04.09.2025**

**Until: 23.11.2027**

The organization was assessed by the Israel Laboratory Accreditation Authority (ISRAC) and found to be worthy of accreditation to the detailed schedule attached.

The schedule is an integral part of this certificate and is numbered with the above certificate number.

Accreditation demonstrates technical competence and operation of an internationally recognized quality management system.

The organization accredited by ISRAC complies with the standards/requirements mentioned above, meets the technical competence requirements and management system requirements that are necessary for it to consistently deliver technically competent results. This accreditation is granted in accordance with the requirements of ISO/IEC 17011:2017, and entails periodic surveillance and reassessment by ISRAC to ensure that the organization continues to comply with the accreditation requirements.

The accreditation is valid provided that the organization continues to meet the criteria as laid down by ISRAC. ISRAC is an EA-MLA (European Accreditation Cooperation Multi-Lateral Agreement) signatory.

This certificate does not constitute an approval in accordance with article 12 of the standard law.

**Date of first accreditation: 24.11. 1997**

**Etty Feller**  
**General Manager**  
**Israel Laboratory Accreditation Authority**

Date of signature 04/09/2025

Page No. 2 of: 9



Name and Address:

Organization name	Calibration Lab. - Elbit Systems EW and SIGINT Elisra Ltd.
Address	29 Hamerkava , Holon, 5885118, Israel
Phone	+972-77-2935053
E-mail (contact person)	<a href="mailto:Gil.Vertman@elbitsystems.com">Gil.Vertman@elbitsystems.com</a>

Site: P or T or M , P-Permanent, T-Temporary, M-Mobile

A permanent (P) or temporary (T) place, or a stationary or mobile (M) facility, at or from which the organization performs activities forming part of its scope of accreditation, starting from sampling to final issuance of a report or certificate and / or quality system activities. A temporary (T) site is a site established under the responsibility of an accredited permanent site. All activities performed at a temporary site are the responsibility of the permanent site. An outdoors work is also considered to be a temporary site. Temporary site will be a site that involves work for special project and the activity will be defined in time (up to 2 years).

Type of Scopes: A- Fixed, C- Flexible scope in analytical tests : Type of matrix, analytes, experimental systems and/or analytical characteristics may be subject to changes, in accordance with the laboratory's approved and documented procedures. For details, please refer to the list of Accredited Tests, available from the laboratory upon request.



Item	Scope Type	Site	Measurand Instrument, Gauge	Range [Including margins] (Does not include margins)	Uncertainty of Measurement <sup>1</sup>	Reference Documents	Remarks
<b>Calibration – Electrical Quantities - DC &amp; LF</b>					<b>כיוול – גדלים חשמליים - זרם ישר ותדר נמוך</b>		
1	A	P	AC Current, Measuring Instruments  זרם חילופין, מכשירי מדידה	1 mA		Metroval software  Directive: Q905-D060-ELS	Datron – 4808 Standard instrument
2	A	P		300 Hz	0.22 $\mu$ A		
3	A	P		1 kHz	0.22 $\mu$ A		
4	A	P		5 kHz	0.25 $\mu$ A		
5	A	P		10 mA			
6	A	P		300 Hz	3.7 $\mu$ A		
7	A	P		1 kHz	2.4 $\mu$ A		
8	A	P		5 kHz	2.2 $\mu$ A		
9	A	P		100 mA			
10	A	P		300 Hz	15 $\mu$ A		
11	A	P		1 kHz	15 $\mu$ A		
12	A	P		5 kHz	24 $\mu$ A		
13	A	P		1 A			
14	A	P		300 Hz	120 $\mu$ A		
			1 kHz	170 $\mu$ A			
			10 A				
			300 Hz	4 mA			
			1 kHz	4 mA			
			5 kHz	4 mA			



Item	Scope Type	Site	Measurand Instrument, Gauge	Range [Including margins] (Does not include margins)	Uncertainty of Measurement <sup>1</sup>	Reference Documents	Remarks	
<b>Calibration – Electrical Quantities - DC &amp; LF</b>					<b>כיוול – גדלים חשמליים - זרם ישר ותדר נמוך</b>			
15	A	P	AC Voltage, Measuring Instruments  מתח חילופין, מכשירי מדידה	100 mV	Metroval software	Datron – 4808 Standard instrument		
16	A	P		1 kHz			16 $\mu$ V	
17	A	P		50 kHz			15 $\mu$ V	
18	A	P		1 V			1 kHz	110 $\mu$ V
19	A	P		20 kHz			50 $\mu$ V	
20	A	P		50 kHz			100 $\mu$ V	
21	A	P		100 kHz			160 $\mu$ V	
22	A	P		300 kHz			400 $\mu$ V	
23	A	P		10 V			1 kHz	500 $\mu$ V
24	A	P		50 kHz			0.6 mV	
25	A	P		100 V			1 kHz	9 mV
26	A	P		750 V			50 kHz	6 mV
27	A	P		1 kHz			0.1 V	
				30 kHz			0.25 V	



Item	Scope Type	Site	Measurand Instrument, Gauge	Range [Including margins] (Does not include margins)	Uncertainty of Measurement <sup>1</sup>	Reference Documents	Remarks	
<b>Calibration – Electrical Quantities - DC &amp; LF</b>					<b>כיוול – גדלים חשמליים - זרם ישר ותדר נמוך</b>			
28	A	P	DC Current, Measuring Instruments	זרם ישר, מכשירי מדידה	100 $\mu$ A	5 nA	Metroval software	
29	A	P			1 mA	30 nA		
30	A	P			10 mA	450 nA		
31	A	P			100 mA	3.9 $\mu$ A		
32	A	P			1 A	58 $\mu$ A		
33	A	P			10 A	3.1 mA		
34	A	P	DC Resistance, Measuring Instruments	התנגדות זרם ישר, מכשירי מדידה	100 $\Omega$	1 m $\Omega$	Metroval software	4 wire connection
35	A	P			1 k $\Omega$	11.7 m $\Omega$		
36	A	P			10 k $\Omega$	0.058 $\Omega$		
37	A	P			100 k $\Omega$	1.2 $\Omega$		
38	A	P			1 M $\Omega$	16.6 $\Omega$		
39	A	P			10 M $\Omega$	310 $\Omega$		
40	A	P	DC Voltage, Measuring Instruments	מתח בזרם ישר, מכשירי מדידה	100 mV	1.8 $\mu$ V	Metroval software	Datron – 4808 Standard instrument
41	A	P			1 V	5.7 $\mu$ V		
42	A	P			10 V	43.5 $\mu$ V		
43	A	P			100 V	0.5 mV		
44	A	P			1 kV	5.85 mV		



Item	Scope Type	Site	Measurand Instrument, Gauge	Range [Including margins] (Does not include margins)	Uncertainty of Measurement <sup>1</sup>	Reference Documents	Remarks
<b>Calibration – Electrical Quantities - RF Frequency, Time</b>					<b>כיוול – גדלים חשמליים - תדר גבוה, זמן</b>		
45	A	P	Frequency, Measuring Instruments תדר, מכשירי מדידה	10 MHz	$1 \times 10^{-11}$	Datum – 9390, User Manual Directive:Q905-D061-ELS	Using Datum GPS Frequency Generator 26 h Integration time
46	A	P	Frequency, Sources תדר, מחוללים	10 MHz to 40 GHz	$1 \times 10^{-10}$	Datum – 9390, User Manual Directive:Q905-D061-ELS	Using Frequency Generator HP 83640A Standard instrument 10 sec. Integration time
47	A	P	RF Power Sensor Calibration Factor, Measuring Instruments מקדם כיוול של גששי הספק תדר גבוה, מכשירי מדידה	10 MHz	1.5 %	Metroval software  Directive: T000-D020-ELS	For N type connectors  The power sensors are calibrated at about 1 mW. The uncertainties are expressed as percentage of Calibration factor value.  CMC value stands for termistor calibrations.  Other sensors may be calibrated at larger uncertainty values.
48	A	P		30 MHz	1.1 %		
49	A	P		50 MHz	1.0 %		
50	A	P		100 MHz	1.1 %		
51	A	P		200 MHz	1.2 %		
52	A	P		300 MHz	1.1 %		
53	A	P		400 MHz	1.3 %		
54	A	P		500 MHz	1.3 %		
55	A	P		1 GHz	1.3 %		
56	A	P		2 GHz	1.2 %		
57	A	P		3 GHz	1.4 %		
58	A	P		4 GHz	1.6 %		
59	A	P		5 GHz	1.9 %		
60	A	P		6 GHz	2.1 %		
61	A	P	7 GHz	2.9 %			



Item	Scope Type	Site	Measurand Instrument, Gauge	Range [Including margins] (Does not include margins)	Uncertainty of Measurement <sup>1</sup>	Reference Documents	Remarks
<b>Calibration – Electrical Quantities - RF Frequency, Time</b>					<b>כיוול – גדלים חשמליים - תדר גבוה, זמן</b>		
62	A	P		8 GHz	4.1 %		
63	A	P		9 GHz	3.6 %		
64	A	P		10 GHz	2.4 %		
65	A	P		11 GHz	2.0 %		
66	A	P		12 GHz	2.0 %		
67	A	P		13 GHz	2.2 %		
68	A	P		14 GHz	2.2 %		
69	A	P		15 GHz	2.8 %		
70	A	P		16 GHz	4.0 %		
71	A	P		17 GHz	2.8 %		
72	A	P		18 GHz	2.6 %		
73	A	P	RF Power Sensor	0.05 GHz	1.4%	Metroval software	For 2.4 mm connectors
74	A	P	Calibration Factor, Measuring Instruments	0.1 GHz	1.2 %	Directive:	Using HP 8487A Standard instrument
75	A	P		0.5 GHz	1.2 %	T000-D020-ELS	The power sensors are calibrated at nominal -5 dBm. The uncertainties are expressed as percentage of Calibration factor.
76	A	P		1 GHz	1.2 %		
77	A	P		2 GHz	1.3 %		
78	A	P		10 GHz	1.6 %		The calibration method and software are developed by the Elisra Electronic System Metrology Laboratory
79	A	P		14 GHz	1.7 %		
80	A	P		18 GHz	2.0 %		



Item	Scope Type	Site	Measurand Instrument, Gauge	Range [Including margins] (Does not include margins)	Uncertainty of Measurement <sup>1</sup>	Reference Documents	Remarks
<b>Calibration – Electrical Quantities - RF Frequency, Time</b>					<b>כיוול – גדלים חשמליים - תדר גבוה, זמן</b>		
81	A	P		20 GHz	2.4 %		
82	A	P		22 GHz	2.3 %		
83	A	P		24 GHz	2.5 %		
84	A	P		26 GHz	2.6 %		
85	A	P		28 GHz	4.2 %		
86	A	P		30 GHz	3.7 %		
87	A	P		32 GHz	3.4 %		
88	A	P		34 GHz	3.5 %		
89	A	P		36 GHz	3.6 %		
90	A	P		38 GHz	3.7 %		
91	A	P		40 GHz	3.5 %		
92	A	P	RF Power, Sources	הספק תדר גבוה 50 MHz	0.75 %	Metroval software	Reference Source 1 mW

<sup>1)</sup> The uncertainty covered by the CMC expressed as the standard measurement uncertainty multiplied by the coverage factor  $k$  such that the coverage probability corresponds to approximately 95 %.