



ISO 15189:2022

מעבדות בדיקה - רפואה

תעודת הסמכה מס' 0083  
מרכז שירותי הדם, מד"א

כתובת אתר ייחוס: שד' הרצל 133, רמלה 71259505

עד יום: 31.12.2027

בתוקף מיום: 21.09.2025

הארגון נבדק ונבחן על ידי הרשות הלאומית להסמכת מעבדות (להלן הרשות) ונמצא ראוי להסמכה בהתאם לנספח פירוט היקף ההסמכה המצורף לתעודה זו, המהווה חלק בלתי נפרד ממנה ומספרו זהה למספר התעודה. הסמכה מצביעה על כשירות מקצועית ותפעול מערכת ניהול איכות בעלת הכרה בינלאומית. הארגון המוסמך על ידי הרשות, עומד בתקנים/ בדרישות המפורטים מעלה. דרישות התקנים הם לכשירות מקצועית ולמערכות ניהול, שהינן הכרחיות למתן תוצאות אמינות. הסמכה זו ניתנה בהתאם לכללי ISO/IEC 17011:2017 לפיהם פועלת הרשות ובמסגרתם מקיימת פיקוח שוטף על הארגון לצורך בחינת תפקודו המתמשך בהתאם לדרישות ההסמכה. ההסמכה תקפה כל עוד הארגון עונה לאמות המידה שנקבעו על ידי הרשות. הרשות חתומה על הסכם הכרה רב צדדי (MLA) מול ארגון European Accreditation Cooperation (EA).

תעודה זו אינה מהווה אישור לפי סעיף 12 לחוק התקנים.

אתי פלר  
מנכ"ל  
הרשות הלאומית להסמכת מעבדות

תאריך הסמכה ראשון: 01.01.2003



הרשות הלאומית להסמכת מעבדות  
Israel Laboratory Accreditation Authority

**Testing Laboratory-Medical**

**ISO 15189:2022**

**Accreditation Certificate No.0083**

**Blood Services Center, MDA**

**Main site address:** Herzl 133 st. Ramla Israel

**Valid from: 21.09.2025**

**Until:31.12.2027**

The organization was assessed by the Israel Laboratory Accreditation Authority (ISRAC) and found to be worthy of accreditation to the detailed schedule attached.

The schedule is an integral part of this certificate and is numbered with the above certificate number.

Accreditation demonstrates technical competence and operation of an internationally recognized quality management system.

The organization accredited by ISRAC complies with the standards/requirements mentioned above, meets the technical competence requirements and management system requirements that are necessary for it to consistently deliver technically competent results. This accreditation is granted in accordance with the requirements of ISO/IEC 17011:2017, and entails periodic surveillance and reassessment by ISRAC to ensure that the organization continues to comply with the accreditation requirements.

The accreditation is valid provided that the organization continues to meet the criteria as laid down by ISRAC. ISRAC is an EA-MLA (European Accreditation Cooperation Multi-Lateral Agreement) signatory.

This certificate does not constitute an approval in accordance with article 12 of the standard law.

**Date of first accreditation: 01.01.2003**

**Etty Feller**  
**General Manager**  
**Israel Laboratory Accreditation Authority**

Date of signature 21/09/2025

Page No. 2 of: 4



Name and Address:  
Organization name **Blood Services Center, MDA**  
Address Herzel 133 st. Ramla Israel  
Phone 972-3-530-0400  
Fax  
E-mail (contact person) [OritF@mda.org.il](mailto:OritF@mda.org.il)

Site: P or T or M , P-Permanent, T-Temporary, M-Mobile

A permanent (P) or temporary (T) place, or a stationary or mobile (M) facility, at or from which the organization performs activities forming part of its scope of accreditation, starting from sampling to final issuance of a report or certificate and / or quality system activities. A temporary (T) site is a site established under the responsibility of an accredited permanent site. All activities performed at a temporary site are the responsibility of the permanent site. An outdoors work is also considered to be a temporary site. Temporary site will be a site that involves work for special project and the activity will be defined in time (up to 2 years).

Type of Scopes: A- Fixed, C- Flexible scope in analytical tests : Type of matrix, analytes, experimental systems and/or analytical characteristics may be subject to changes, in accordance with the laboratory's approved and documented procedures. For details, please refer to the list of Accredited Tests, available from the laboratory upon request.



Item	Scope Type	Site	Types of Test / Properties Measured	Materials / Samples Tested	Test principle / Characteristics	Equipment used / Work station	Opinion and Interpretation	Remarks	
<b>Group of products: Health, Medicine - Blood Bank</b>					<b>משפחת מוצרים: בריאות, רפואה - בנק דם</b>				
<b>Blood Component Preparation, CBC</b>					<b>הכנת מרכיבי דם, ספירת דם</b>				
1	A	P	Determination of factor VIII	קביעת פקטור VIII	Plasma, Cryoprecipitate	פלסמה, פלסמה, Cryoprecipitate	Clotting time	ACL TOP 350 coagulation system	---
2	A	P	Hct (Hematocrit)	Hct (Hematocrit)	Whole blood	דם מלא	Calculation from RBC count and mean red cell volume	UniCel DxH 600 Coulter Cellular Analysis System	---
3	A	P	Hgb (Hemoglobin)	המוגלובין	Whole blood	דם מלא	Photometric measurement	UniCel DxH 600 Coulter Cellular Analysis System	---
4	A	P	PLT(Platelets)	PLT (Platelets)	Whole blood	דם מלא	Coulter principle	UniCel DxH 600 Coulter Cellular Analysis System	---
5	A	P	RBC (Red Blood Cells)	תאי דם אדומים	Whole blood	דם מלא	Coulter principle	UniCel DxH 600 Coulter Cellular Analysis System	---
6	A	P	WBC (White Blood Cells)	תאי דם לבנים	Whole blood	דם מלא	Coulter principle	UniCel DxH 600 Coulter Cellular Analysis System	---