



ISO/IEC 17025:2017

מעבדות כיוול

תעודת הסמכה מס' 0094 חברת החשמל מעבדת תקנים

כתובת אתר ייחוס: המדע 6, ראשל"צ, 7559909

עד יום: 11.02.2028

בתוקף מיום: 02.02.2025

הארגון נבדק ונבחן על ידי הרשות הלאומית להסמכת מעבדות (להלן הרשות) ונמצא ראוי להסמכה בהתאם לנספח פירוט היקף ההסמכה המצורף לתעודה זו, המהווה חלק בלתי נפרד ממנה ומספרו זהה למספר התעודה. הסמכה מצביעה על כשירות מקצועית ותפעול מערכת ניהול איכות בעלת הכרה בינלאומית. הארגון המוסמך על ידי הרשות, עומד בתקנים / דרישות המפורטים מעלה. דרישות התקנים הם לכשירות מקצועית ולמערכות ניהול, שהינן הכרחיות למתן תוצאות אמינות. הסמכה זו ניתנה בהתאם לכללי תקן ISO/IEC 17011:2017 לפיהם פועלת הרשות ובמסגרתם מקיימת פיקוח שוטף על הארגון לצורך בחינת תפקודו המתמשך בהתאם לדרישות ההסמכה. ההסמכה תקפה כל עוד הארגון עונה לאמות המידה שנקבעו על ידי הרשות. הרשות חתומה על הסכם הכרה זו צדדי (BLA) מול ארגון (EA) European co-operation for Accreditation.

תעודה זו אינה מהווה אישור לפי סעיף 12 לחוק התקנים.

אתי פלר
מנכ"ל
הרשות הלאומית להסמכת מעבדות

תאריך הסמכה ראשון: 12.02.2003



הרשות הלאומית להסמכת מעבדות
Israel Laboratory Accreditation Authority

Calibration Laboratories

ISO/IEC 17025:2017

Accreditation Certificate No. 0094

Israel Electric Corp Standard Laboratory

Main site address: 6 Hamada st. Rishon LeTzion, 7559909, Israel

Valid from: 02.02.2025

Until: 11.02.2028

The organization was assessed by the Israel Laboratory Accreditation Authority (ISRAC) and found to be worthy of accreditation to the detailed schedule attached.

The schedule is an integral part of this certificate and is numbered with the above certificate number.

Accreditation demonstrates technical competence and operation of an internationally recognized quality management system.

The organization accredited by ISRAC complies with the standards / requirements mentioned above, meets the technical competence requirements and management system requirements that are necessary for it to consistently deliver technically competent results. This accreditation is granted in accordance with the requirements of the standard ISO/IEC 17011:2017, and entails periodic surveillance and reassessment by ISRAC to ensure that the organization continues to comply with the accreditation requirements.

The accreditation is valid provided that the organization continues to meet the criteria as laid down by ISRAC. ISRAC is an EA-BLA (European co-operation for Accreditation Bi-Lateral Agreement) signatory.

This certificate does not constitute an approval in accordance with article 12 of the standard law.

Date of first accreditation: 12.02.2003

**Etty Feller
General Manager
Israel Laboratory Accreditation Authority**

Date of signature 02/02/2025

Page No. 2 of: 20



Name and Address:

Organization name Israel Electric Corp Standard Laboratory 6 HaMada st., Rishon LeTzion, 7559909,
Address Israel

Phone +972-076-8647850
Fax +972-072-3417851
E--mail (contact person) itshak.gueta@iec.co.il

- *P* Israel Electric Corp. Ltd Meter Test Station
Marketing Division Central Metering Unit
Standard Laboratory
- *PI* H.V Laboratory-Beit Dagan Store, +972-3-960-2241
Meter Testing Department-Central Metering Unit,
Marketing Division, Israel Electric Corporation Ltd.

As declared by the Organization, following is a list of the permanent sites and the phones numbers at which accredited activities are performed

Site: P or T or M , P-Permanent, T-Temporary, M-Mobile

A permanent (P) or temporary (T) place, or a stationary or mobile (M) facility, at or from which the organization performs activities forming part of its scope of accreditation, starting from sampling to final issuance of a report or certificate and / or quality system activities. A temporary (T) site is a site established under the responsibility of an accredited permanent site. All activities performed at a temporary site are the responsibility of the permanent site. An outdoors work is also considered to be a temporary site. Temporary site will be a site that involves work for special project and the activity will be defined in time (up to 2 years).

Type of Scopes: A- Fixed, C- Flexible scope in analytical tests : Type of matrix, analytes, experimental systems and/or analytical characteristics may be subject to changes, in accordance with the laboratory's approved and documented procedures. For details, please refer to the list of Accredited Tests, available from the laboratory upon request.



Item	Scope Type	Site	Measurand, Instrument, Gauge		Range [Including margins] (Does not include margins)	Uncertainty of Measurement ¹	Reference Documents	Remarks
Calibration – Electrical Quantities - DC and LF						כיוול – גדלים חשמליים - זרם ישר ותדר נמוך		
1	A	P	DC Voltage, Sources	מחוללי מתח בזרם ישר	[0 V to 0.1 V]	2 μ V	Manufacturer Instructions	Digital Multimeters: Fluke 8588A; Fluke 8508A
2	A	P			(0.1 V to 0.2 V)	2.7 μ V		
3	A	P			(0.2 V to 1.0 V)	14 μ V		
4	A	P			(1.0 V to 2.0 V)	19 μ V		
5	A	P			(2.0 V to 10 V)	120 μ V		
6	A	P			(10 V to 20 V)	180 μ V		
7	A	P			(20 V to 100 V)	1.4 mV		
8	A	P			(100 V to 200 V)	2.3 mV		
9	A	P			(200 V to 1000 V)	14 mV		
10	A	P	DC Voltage Reference Voltage Sources	מתח בזרם ישר, מקורות מתח ישר לייחוס	1.018 V	2.2 μ V	Manufacturer Instructions	Standard Cells Fluke 7001 Digital Multimeters: Fluke 8588A; Fluke 8508A
11	A	P			10 V	20 μ V		
12	A	P	DC Voltage, Measuring Instruments	מכשירי מדידת מתח בזרם ישר	[0 V to 0.33 V]	16 μ V	Manufacturer Instructions	Calibrator Fluke 5520A
13	A	P			(0.33 V to 3.3 V)	160 μ V		
14	A	P			(3.3 V to 33 V)	1.6 mV		
15	A	P			(33 V to 330 V)	21 mV		
16	A	P			(330 V to 1020 V)	21 mV		
17	A	P	DC Current, Sources	זרם ישר, מחוללים	[0 mA to 0.2 mA]	5.4 nA	Manufacturer Instructions	Digital Multimeters Fluke 8588A; Fluke 8508A
18	A	P			(0.2 mA to 2.0 mA)	54 nA		
19	A	P			(2.0 mA to 20 mA)	0.57 μ A		



Item	Scope Type	Site	Measurand, Instrument, Gauge		Range [Including margins] (Does not include margins)	Uncertainty of Measurement ¹	Reference Documents	Remarks
Calibration – Electrical Quantities - DC and LF						כיוול – גדלים חשמליים - זרם ישר ותדר נמוך		
20	A	P			(20 mA to 200 mA]	17 μ A		
21	A	P			(200 mA to 1.0 A]	250 μ A		
22	A	P			(1.0 A to 2.0 A]	0.58 mA		
23	A	P			(2.0 A to 20 A]	13 mA		
24	A	P	DC Current, Measuring Instruments	זרם ישר, מכשירי מדידה	[0 mA to 0.32 mA]	0.7 μ A	Manufacturer Instructions	Calibrators Fluke 5520A;
25	A	P			(0.32 mA to 3.3 mA]	1.2 μ A		
26	A	P			(3.3 mA to 33 mA]	21 μ A		
27	A	P			(33 mA to 330 mA]	310 μ A		
28	A	P			(330 mA to 2.1 A]	1.2 mA		
29	A	P			(3.2 A to 11 A]	12 mA		
30	A	P	AC Voltage, Sources	מתח חילופין מחוללים	[40 Hz to 70 Hz]		Manufacturer Instructions	Digital Multimeters Fluke 8588A; Fluke 8508A
31	A	P			[0.00 V to 0.1 V]	0.053 mV		
32	A	P			(0.1 V to 0.2 V]	0.072 mV		
33	A	P			(0.2 V to 1.0 V]	0.2 mV		
34	A	P			(1.0 V to 2.0 V]	0.41 mV		
35	A	P			(2.0 V to 10 V]	3.8 mV		
36	A	P			(10 V to 20 V]	4.9 mV		
37	A	P			(20 V to 60 V]	6.0 mV		
				(60 V to 120 V]	11 mV		Three Phase Comparator Zera COM 303-1	



Item	Scope Type	Site	Measurand, Instrument, Gauge	Range [Including margins] (Does not include margins)	Uncertainty of Measurement ¹	Reference Documents	Remarks
Calibration – Electrical Quantities - DC and LF					כיוול – גדלים חשמליים - זרם ישר ותדר נמוך		
38	A	P		(120 V to 240 V)	22 mV		
39	A	P		(240 V to 480 V)	44 mV		Digital Multimeter Fluke 8508A
40	A	P	AC Voltage. Measuring Instruments	מתח חילופין מכשירי מדידה	[40 Hz to 70 Hz] [0.00 V to 0.01 V]	Manufacturer Instructions	Calibrators Fluke 5520A;
41	A	P		(0.01 V to 0.033 V)	160 μV		
42	A	P		(0.033 V to 0.32 V)	600 μV		
43	A	P		(0.32 V to 3.2 V)	5.4 mV		
44	A	P		(3.2 V to 33 V)	40 mV		
45	A	P		(33 V to 330 V)	280 mV		
46	A	P		(330 V to 1020 V)	1 V		
47	A	P	AC Voltage. Sources	מתח חילופין מחוללים	[70 Hz to 1 KHz] [0.00 V to 0.1 V]	Manufacturer Instructions	Digital Multimeters Fluke 8588A; Fluke 8508A
48	A	P		(0.1 V to 0.2 V)	0.075 mV		
49	A	P		(0.2 V to 1.0 V)	0.18 mV		
50	A	P		(1.0 V to 2.0 V)	0.46 mV		
51	A	P		(2.0 V to 10 V)	3.8 mV		
52	A	P		(10 V to 20 V)	5.4 mV		
53	A	P		(20 V to 100 V)	24 mV		
54	A	P		(100 V to 200 V)	46 mV		



Item	Scope Type	Site	Measurand, Instrument, Gauge	Range [Including margins] (Does not include margins)	Uncertainty of Measurement ¹	Reference Documents	Remarks
Calibration – Electrical Quantities - DC and LF					כיוול – גדלים חשמליים - זרם ישר ותדר נמוך		
55	A	P		(200V to 1000V]	240 mV		
56	A	P	AC Current Sources	זרם חילופין מחוללים [40 Hz to 70 Hz]	0.073 μ A	Manufacturer Instructions	Digital Multimeters Fluke 8588A; Fluke 8508A
57	A	P		(100 μ A to 200 μ A]	0.17 μ A		
58	A	P		(200 μ A to 1.0 mA]	0.064 μ A		
59	A	P		(1.0 mA to 2.0 mA]	1.2 μ A		
60	A	P		(2.0 mA to 5.0 mA]	1.3 μ A		Three Phase Comparator Zera COM 303-1
61	A	P		(5.0 mA to 10 mA]	1.6 μ A		
62	A	P		(10 mA to 20 mA]	4.0 μ A		
63	A	P		(20 mA to 50 mA]	8.0 μ A		
64	A	P		(50 mA to 100 mA]	10 μ A		
65	A	P		(100mA to 200mA]	21 μ A		
66	A	P		(200mA to 500mA]	50 μ A		
67	A	P		(0.5 A to 1.0 A]	0.1 mA		
68	A	P		(1.0 A to 2.0 A]	0.21 mA		
69	A	P		(2.0 A to 5.0 A]]	0.5 mA		
70	A	P		(5.0 A to 10 A]	1.0 mA		
71	A	P		(10 A to 20 A]	4.0 mA		
72	A	P		(20 A to 50 A]	10 mA		



Item	Scope Type	Site	Measurand, Instrument, Gauge	Range [Including margins] (Does not include margins)	Uncertainty of Measurement ¹	Reference Documents	Remarks	
Calibration – Electrical Quantities - DC and LF					כיוול – גדלים חשמליים - זרם ישר ותדר נמוך			
73	A	P		(50 A to 100 A]	18 mA			
74	A	P		(100 A to 160 A]	43 mA			
75	A	P		(160 A to 500 A]	157 mA		Three Phase Comparator	
76	A	P		(500 A to 1000 A]	197 mA		Zera COM 303-1 Current Comparator Tettex 4761	
77	A	P	AC Current Measuring Instruments	זרם חילופין מכשירי מדידה [40 Hz to 70 Hz]	[0.0 μA to 33 μA]	0.92 μA	Manufacturer Instructions	Calibrators Fluke 5520A;
78	A	P		(33 μA to 0.32 mA]	1.3 μA			
79	A	P		0.32mA to 3.2mA]	5.2 μA			
80	A	P		(3.2 mA to 33 mA]	56 μA			
81	A	P		(33 mA to 330 mA]	560 μA			
82	A	P		(0.33 A to 2.2 A]	4.2 mA			
83	A	P		(2.2 A to 11 A]	14 mA			
84	A	P		(11 A to 20 A]	70 mA			
85	A	P		(20 A to 32 A]	110 mA			
86	A	P		(32 A to 160 A]	560 mA			
87	A	P		(160 A to 200 A]	750 mA			
88	A	P		(200 A to 1000 A]	3.9 A			



Item	Scope Type	Site	Measurand, Instrument, Gauge	Range [Including margins] (Does not include margins)	Uncertainty of Measurement ¹	Reference Documents	Remarks		
Calibration – Electrical Quantities - DC and LF					כיוול – גדלים חשמליים - זרם ישר ותדר נמוך				
89	A	P	AC Current Sources	זרם חילופין מחוללים [70 Hz to 1 KHz] [0.0μA to 0.1mA]	0.073 μA	Manufacturer Instructions	Digital Multimeters Fluke 8588A Fluke 8508A		
90	A	P						(0.1 mA to 0.2mA)	0.2 μA
91	A	P						(0.2 mA to 1.0mA)	0.64 μA
92	A	P						(1.0 mA to 2.0mA)	1.4 μA
93	A	P						(2.0 mA to 10 mA)	6.4 μA
94	A	P						(10 mA to 20 mA)	14 μA
95	A	P						(20 mA to 100mA)	45 μA
96	A	P						(100mA to 200mA)	170 μA
97	A	P						(0.2 A to 1.0 A)	0.95 mA
98	A	P						(1.0 A to 2.0 A)	1.9 mA
99	A	P						(2.0 A to 20 A)	25 mA
100	A	P	DC resistance, Reference Resistance	התנגדות זרם ישר, התנגדויות לייחוס (0.0 Ω to 0.02 Ω)	13 μΩ	Manufacturer Instructions	Micro-Ohmmeter Cambridge 510A; Digital Multimeters Fluke 8588A; Fluke 8508A		
101	A	P						(0.02 Ω to 2.0 Ω)	77 μΩ
102	A	P						(2.0 Ω to 20 Ω)	0.37 mΩ
103	A	P						(20 Ω to 100 Ω)	1.9 mΩ
104	A	P						(100 Ω to 200 Ω)	3.5 mΩ
105	A	P						(200 Ω to 1.0 kΩ)	1.9 mΩ
106	A	P						(1.0 kΩ to 2.0 kΩ)	35 mΩ
107	A	P						(2.0 kΩ to 10 kΩ)	0.17 Ω



Item	Scope Type	Site	Measurand, Instrument, Gauge	Range [Including margins] (Does not include margins)	Uncertainty of Measurement ¹	Reference Documents	Remarks
Calibration – Electrical Quantities - DC and LF					כיוול – גדלים חשמליים - זרם ישר ותדר נמוך		
108	A	P		(10 kΩ to 20 kΩ]	0.30 Ω		
109	A	P		(20 kΩ to 100 kΩ]	1.9 Ω		
110	A	P		(100 kΩ to 200 kΩ]	3.5 Ω		
111	A	P		(200 kΩ to 1.0 MΩ]	31 Ω		
112	A	P		(1.0 MΩ to 2.0 MΩ]	50 Ω		
113	A	P		(2.0 MΩ to 10 MΩ]	0.64 kΩ		
114	A	P		(10 MΩ to 20 MΩ]	0.93 kΩ		
115	A	P		(20 MΩ to 200 MΩ]	63 kΩ		Digital Multimeter
116	A	P		(200 MΩ to 1 GΩ]	5.1 MΩ		Fluke 8508A
117	A	P	DC Resistance, Measuring Instruments	(0.0 Ω to 11.0 Ω]	14 mΩ	Manufacturer Instructions	Calibrators
118	A	P		(11.0 Ω to 33.0 Ω]	26 mΩ		Fluke 5520A;
119	A	P		(33 Ω to 110 Ω]	40 mΩ		
120	A	P		(110 Ω to 330 Ω]	100 mΩ		
121	A	P		(330 Ω to 1.1 kΩ]	300 mΩ		
122	A	P		(1.1 kΩ to 3.3 kΩ]	1.0 Ω		
123	A	P		(3.3 kΩ to 11 kΩ]	3.0 Ω		
124	A	P		(11 kΩ to 33 kΩ]	7.0 Ω		
125	A	P		(33 kΩ to 110 kΩ]	28 Ω		
126	A	P		(110 kΩ to 330 kΩ]	85 Ω		



Item	Scope Type	Site	Measurand, Instrument, Gauge	Range [Including margins] (Does not include margins)	Uncertainty of Measurement ¹	Reference Documents	Remarks
Calibration – Electrical Quantities - DC and LF					כיוול – גדלים חשמליים - זרם ישר ותדר נמוך		
127	A	P		(330 kΩ to 1.1 MΩ]	350 Ω		
128	A	P		(1.1 MΩ to 3.3 MΩ]	1.2 kΩ		
129	A	P		(3.3 MΩ to 11 MΩ]	12 kΩ		
130	A	P		(11 MΩ to 33 MΩ]	100 kΩ		
131	A	P		(33 MΩ to 110 MΩ]	920 kΩ		
132	A	P		(110 MΩ to 330 MΩ]	6.0 MΩ		
133	A	P	Phase angle, Wattmeter, Power source	זווית מופע, מד הספק, מקור הספק	[0° to 360°]	0.034°	Manufacturer Instructions Three Phase Comparator Zera COM 303-1
134	A	P1	Voltage ratio Voltage measuring transformer Primary: 6 kV to 40 kV Secondary: 110 V or 110/√3 V	יחס השנאה של שנאי מתח	1/300 to 1/120	2·10 ⁻⁴ V/V	IEC 61869-3 Reference V.T. EPRO NVOD 30. Measuring Bridge Zera WM-303U 1' (minute) stands for 0.291 mrad
135	A	P1	Phase displacement Voltage measuring transformer Primary: 6 kV to 40 kV Secondary: 110 V or 110/√3 V	הזזת מופע של שנאי מתח	0' to 100'	2.0'	
136	A	P1	Current ratio Current measuring transformer Primary: 0.05 A to 1000 A Secondary: 5 A	יחס השנאה של שנאי זרם	1/200 to 1/1	1.1·10 ⁻⁴ A/A	IEC 61869-1 IEC 61869-2 Current comparator TETTEX 4761. Measuring Bridge Zera WM-303I 1' (minute) stands for 0.291 mrad
137	A	P1	Phase displacement Current measuring transformer Primary: 0.05A to 1000A Secondary: 5A	הזזת מופע של שנאי זרם	0' to 100'	0.3'	



Item	Scope Type	Site	Measurand, Instrument, Gauge	Range [Including margins] (Does not include margins)	Uncertainty of Measurement ¹	Reference Documents	Remarks
Calibration – Electrical Quantities - DC and LF					כיוול – גדלים חשמליים - זרם ישר ותדר נמוך		
138	A	P	Frequency, Sources	תדר, מחוללים [0.5 Hz to 10.0 MHz]	1.5 Hz / MHz	Manufacturer Instructions	Timer/ Counter Pendulum CNT 80
139	A	P	Frequency. Counters	תדר, מוני תדר [0.5 Hz to 10.0 MHz]	2.5 Hz / MHz	Manufacturer Instructions	Waveform Generator Agilent 33500B
140	A	P	Active Energy, in single phase 50 Hz network,	63.5 V P. Factor 0.5	200 μJ/J	IEC 60736 IEC 60514 IEC 61358 IEC 60687 IEC 62052-11 IEC 62053-11 IEC 62053-21 IEC 62053-22 IEC 62053-23 IEC 60688 IEC 60145	Reference instrument COM 303 standard instrument Voltage measured phase to neutral
141	A	P		5 A P. Factor 1			
142	A	P		10 A			
143	A	P		230 V P. Factor 0.5			
144	A	P		2.5 A			
145	A	P		5 A			
146	A	P		10 A			
147	A	P		20 A			
148	A	P		50 A			
149	A	P		100 A			
150	A	P		P. Factor 1			
151	A	P		0.05 A			
152	A	P		0.1 A			
153	A	P		0.2 A			
154	A	P	0.25 A				
	A	P	0.5 A				
	A	P	1 A				



Item	Scope Type	Site	Measurand, Instrument, Gauge	Range [Including margins] (Does not include margins)	Uncertainty of Measurement ¹	Reference Documents	Remarks
Calibration – Electrical Quantities - DC and LF					כיוול – גדלים חשמליים - זרם ישר ותדר נמוך		
155	A	P		2 A	150 μJ/J		
156	A	P		2.5 A	150 μJ/J		
157	A	P		5 A	150 μJ/J		
158	A	P		10 A	150 μJ/J		
159	A	P		20 A	170 μJ/J		
160	A	P		50 A	170 μJ/J		
161	A	P		100 A	170 μJ/J		
162	A	P	Active Energy, in three phase 50 Hz network, 3 wires, 2 elements	110 V P. Factor 0.5 0.01 A	220 μJ/J	IEC 60736 IEC 60514 IEC 61358 IEC 60687 IEC 62052-11 IEC 62053-11 IEC 62053-21 IEC 62053-22 IEC 62053-23	Reference instrument COM 303 of Zera standard instrument. Voltage measured phase to phase
163	A	P		0.02 A	220 μJ/J		
164	A	P		0.05 A	220 μJ/J		
165	A	P		0.1 A	200 μJ/J		
166	A	P		0.2 A	200 μJ/J		
167	A	P		0.5 A	200 μJ/J		
168	A	P		1 A	200 μJ/J		
169	A	P		2 A	200 μJ/J		
170	A	P		5 A	200 μJ/J		
171	A	P		P. Factor 1 0.01 A	170 μJ/J		
172	A	P		0.02 A	170 μJ/J		
173	A	P		0.05 A	170 μJ/J		
174	A	P		0.1 A	150 μJ/J		
175	A	P		0.2 A	150 μJ/J		



Item	Scope Type	Site	Measurand, Instrument, Gauge	Range [Including margins] (Does not include margins)	Uncertainty of Measurement ¹	Reference Documents	Remarks
Calibration – Electrical Quantities - DC and LF					כיוול – גדלים חשמליים - זרם ישר ותדר נמוך		
176	A	P		0.5 A	150 μ N/J		
177	A	P		1 A	150 μ N/J		
178	A	P		2 A	150 μ N/J		
179	A	P		5 A	150 μ N/J		
180	A	P		10 A	150 μ N/J		
181	A	P		20 A	170 μ N/J		
182	A	P	Active energy, in three phase 50 Hz network, 4 wires, 3 elements	63.5 V P. Factor 0.5 0.05 A	220 μ N/J		Reference instrument COM 303 of Zera standard instrument. Voltage measured phase to neutral
183	A	P		0.1 A	200 μ N/J		
184	A	P		0.2 A	200 μ N/J		
185	A	P		0.5 A	200 μ N/J		
186	A	P		1 A	200 μ N/J		
187	A	P		2 A	200 μ N/J		
188	A	P		5 A	200 μ N/J		
189	A	P		10 A	200 μ N/J		
190	A	P		P. Factor 1 0.05 A	170 μ N/J		
191	A	P		0.1 A	150 μ N/J		
192	A	P		0.2 A	150 μ N/J		
193	A	P		0.5 A	150 μ N/J		
194	A	P		1 A	150 μ N/J		
195	A	P		2 A	150 μ N/J		
196	A	P		5 A	150 μ N/J		
197	A	P		10 A	150 μ N/J		



Item	Scope Type	Site	Measurand, Instrument, Gauge	Range [Including margins] (Does not include margins)	Uncertainty of Measurement ¹	Reference Documents	Remarks
Calibration – Electrical Quantities - DC and LF					כיוול – גדלים חשמליים - זרם ישר ותדר נמוך		
				230 V			
				P. Factor 0.5			
198	A	P		0.01 A	220 μJ/J		
199	A	P		0.02 A	220 μJ/J		
200	A	P		0.05 A	220 μJ/J		
201	A	P		0.1 A	200 μJ/J		
202	A	P		0.2 A	200 μJ/J		
203	A	P		0.5 A	200 μJ/J		
204	A	P		1 A	200 μJ/J		
205	A	P		2 A	200 μJ/J		
206	A	P		5 A	200 μJ/J		
207	A	P		10 A	200 μJ/J		
208	A	P		20 A	220 μJ/J		
209	A	P		50 A	220 μJ/J		
210	A	P		100 A	220 μJ/J		
				P. Factor 1			
211	A	P		0.01 A	170 μJ/J		
212	A	P		0.02 A	170 μJ/J		
213	A	P		0.05 A	170 μJ/J		
214	A	P		0.1 A	150 μJ/J		



Item	Scope Type	Site	Measurand, Instrument, Gauge	Range [Including margins] (Does not include margins)	Uncertainty of Measurement ¹	Reference Documents	Remarks
Calibration – Electrical Quantities - DC and LF					כיוול – גדלים חשמליים - זרם ישר ותדר נמוך		
215	A	P		0.2 A	150 µN/J		
216	A	P		0.5 A	150 µN/J		
217	A	P		1 A	150 µN/J		
218	A	P		2 A	150 µN/J		
219	A	P		5 A	150 µN/J		
220	A	P		10 A	150 µN/J		
221	A	P		20 A	170 µN/J		
222	A	P		50 A	170 µN/J		
223	A	P		100 A	170 µN/J		
224	A	P	Reactive energy, in three phase 50 Hz network, 3 wires, 2 elements	110 V P. Factor 0.5 0.05 A	300 µN/J		Reference instrument COM 303 of Zera standard instrument. Voltage measured phase to phase
225	A	P		0.1 A	280 µN/J		
226	A	P		0.2 A	280 µN/J		
227	A	P		0.5 A	280 µN/J		
228	A	P		1 A	280 µN/J		
229	A	P		2 A	280 µN/J		
230	A	P		5 A	280 µN/J		
231	A	P		10 A	280 µN/J		



Item	Scope Type	Site	Measurand, Instrument, Gauge	Range [Including margins] (Does not include margins)	Uncertainty of Measurement ¹	Reference Documents	Remarks
Calibration – Electrical Quantities - DC and LF					כיוול – גדלים חשמליים - זרם ישר ותדר נמוך		
232	A	P		P. Factor 1 0.01 A	240 µN/J		
233	A	P		0.02 A	240 µN/J		
234	A	P		0.05 A	240 µN/J		
235	A	P		0.1 A	230 µN/J		
236	A	P		0.2 A	230 µN/J		
237	A	P		0.5 A	230 µN/J		
238	A	P		1 A	230 µN/J		
239	A	P		2 A	230 µN/J		
240	A	P		5 A	230 µN/J		
241	A	P		10 A	230 µN/J		
242	A	P	Reactive energy, in three phase 50 Hz network, 4 wires, 3 elements	63,5 V P. Factor 0.5 0.05 A	300 µN/J		Reference instrument COM 303 of Zera standard instrument. Voltage measured phase to neutral
243	A	P		0.1 A	280 µN/J		
244	A	P		0.2 A	280 µN/J		
245	A	P		0.5 A	280 µN/J		
246	A	P		1 A	280 µN/J		
247	A	P		2 A	280 µN/J		



Item	Scope Type	Site	Measurand, Instrument, Gauge	Range [Including margins] (Does not include margins)	Uncertainty of Measurement ¹	Reference Documents	Remarks
Calibration – Electrical Quantities - DC and LF					כיוול – גדלים חשמליים - זרם ישר ותדר נמוך		
248	A	P		5 A	280 μ N/J		
249	A	P		10 A	280 μ N/J		
				P. Factor 1			
250	A	P		0.01 A	240 μ N/J		
251	A	P		0.02 A	240 μ N/J		
252	A	P		0.05 A	240 μ N/J		
253	A	P		0.1 A	230 μ N/J		
254	A	P		0.2 A	230 μ N/J		
255	A	P		0.5 A	230 μ N/J		
256	A	P		1 A	230 μ N/J		
257	A	P		2 A	230 μ N/J		
258	A	P		5 A	230 μ N/J		
259	A	P		10 A	230 μ N/J		
				230 V			
				P. Factor 0.5			
260	A	P		0.05 A	300 μ N/J		
261	A	P		0.1 A	280 μ N/J		
262	A	P		0.2 A	280 μ N/J		
263	A	P		0.5 A	280 μ N/J		



Item	Scope Type	Site	Measurand, Instrument, Gauge	Range [Including margins] (Does not include margins)	Uncertainty of Measurement ¹	Reference Documents	Remarks
Calibration – Electrical Quantities - DC and LF					כיוול – גדלים חשמליים - זרם ישר ותדר נמוך		
264	A	P		1 A	280 μN/J		
265	A	P		2 A	280 μN/J		
266	A	P		5 A	280 μN/J		
267	A	P		10 A	280 μN/J		
				P. Factor 1			
268	A	P		0.01 A	240 μN/J		
269	A	P		0.02 A	240 μN/J		
270	A	P		0.05 A	240 μN/J		
271	A	P		0.1 A	230 μN/J		
272	A	P		0.2 A	230 μN/J		
273	A	P		0.5 A	230 μN/J		
274	A	P		1 A	230 μN/J		
275	A	P		2 A	230 μN/J		
276	A	P		5 A	230 μN/J		
277	A	P		10 A	230 μN/J		



Item	Scope Type	Site	Measurand Instrument, Gauge	Range [Including margins] (Does not include margins)	Uncertainty of Measurement ¹	Reference Documents	Remarks
Calibration – Physical Quantities - Humidity					כיוול – גדלים פיזיקליים - לחות		
278	A	P	Relative Humidity, Hygrometers and humidity transducers, Ambient Humidity measurement	לחות יחסית, מדי לחות ומתמרים, מדידת לחות סביבה	23 °C [30 % rh to 95 % rh]	IEC 60068 A Guide to the Measurement of humidity- NPL	Reference instrument ALMEMO 2590-2A
279	A	P	Humidity measurement	50 °C [10 % rh to 95 % rh]	5.0 % rh 6.0 % rh		

Item	Scope Type	Site	Measurand Instrument, Gauge	Range [Including margins] (Does not include margins)	Uncertainty of Measurement ¹	Reference Documents	Remarks
Calibration – Physical Quantities - Temperature					כיוול – גדלים פיזיקליים - טמפרטורה		
280	A	P	Temperature, Temperature Meters and Transducers	טמפרטורה מדי טמפרטורה ומתמרים	[0 °C to 60 °C]	IEC 60068	Reference instrument ALMEMO 2590-2A

¹) The uncertainty covered by the CMC expressed as the standard measurement uncertainty multiplied by the coverage factor k such that the coverage probability corresponds to approximately 95 %.