



ISO/IEC 17025:2017

מעבדות כיוול

תעודת הסמכה מס' 0094
חברת החשמל לישראל בע"מ; אגף הנדסת רשת; מונים ארצי-
מחלקת מבדקות מונים; מעבדת תקנים

כתובת אתר ייחוס: לח"י 17, בני ברק, 5120026

עד יום: 11.02.2022

בתוקף מיום: 09.02.2020

הארגון נבדק ונבחן על ידי הרשות הלאומית להסמכת מעבדות (להלן הרשות) ונמצא ראוי להסמכה בהתאם לנספח פירוט היקף ההסמכה המצורף לתעודה זו, המהווה חלק בלתי נפרד ממנה ומספרו זהה למספר התעודה. הסמכה מצביעה על כשירות מקצועית ותפעול מערכת ניהול איכות בעלת הכרה בינלאומית. הארגון המוסמך על ידי הרשות, עומד בתקנים/ בדרישות המפורטים למעלה. דרישות התקנים הם לכשירות מקצועית ולמערכות ניהול, שהינן הכרחיות למתן תוצאות אמינות. הסמכה זו ניתנה בהתאם לכללי ISO/IEC 17011:2017 לפיהם פועלת הרשות ובמסגרתם מקיימת פיקוח שוטף על הארגון לצורך בחינת תפקודו המתמשך בהתאם לדרישות ההסמכה. ההסמכה תקפה כל עוד הארגון עונה לאמות המידה שנקבעו על ידי הרשות.

תעודה זו אינה מהווה אישור לפי סעיף 12 לחוק התקנים.

תאריך הסמכה ראשון: 12.02.2003

אתי פלר
מנכ"ל
הרשות הלאומית להסמכת מעבדות



הרשות הלאומית להסמכת מעבדות
Israel Laboratory Accreditation Authority

Calibration Laboratories

ISO/IEC 17025:2017

Accreditation Certificate No. 94

Standards Laboratory; Meter Testing Department - Central Metering Unit; Network Engineering Division; Israel Electric Corporation Ltd.

Main site address: 17 Lechi st., Bnei-Brak ,5120026, Israel

Valid from: 09.02.2020

Until: 11.02.2022

The organization was assessed by the Israel Laboratory Accreditation Authority (ISRAC) and found to be worthy of accreditation to the detailed schedule attached.

The schedule is an integral part of this certificate and is numbered with the above certificate number.

Accreditation demonstrates technical competence and operation of an internationally recognized quality management system.

The organization accredited by ISRAC complies with the standards/requirements mentioned above, meets the technical competence requirements and management system requirements that are necessary for it to consistently deliver technically competent results. This accreditation is granted in accordance with the requirements of ISO/IEC 17011:2017, and entails periodic surveillance and reassessment by ISRAC to ensure that the organization continues to comply with the accreditation requirements.

The accreditation is valid provided that the organization continues to meet the criteria as laid down by ISRAC.

This certificate does not constitute an approval in accordance with article 12 of the standard law.

Date of first accreditation: 12.02.2003

**Etty Feller
General Manager
Israel Laboratory Accreditation Authority**

Date of signature 03/06/2020

Page No. 2 of: 21



Name and Address:

Laboratory name

Standards Laboratory, Meter Testing Department - Central Metering Unit,
Network Engineering Division, Israel Electric Corporation Ltd.

Address

17 Lechi St.
Bnei-Brak 5120026, Israel

Phone

+972-076-8647850

Fax

+972-072-3417851

E-Mail

itzikge@iec.co.il

- *P* Israel Electric Corp. Ltd Meter Test Station
Marketing Division Central Metering Unit
Standard Laboratory
- *PI* H.V Laboratory-Beit Dagan Store, +972-3-960-2241
Meter Testing Department-Central Metering Unit,
Marketing Division, Israel Electric Corporation Ltd.

As declared by the Organization, following is a list of the permanent sites and the phones numbers at which accredited activities are performed

Site: P or T or M , P-Permanent, T-Temporary, M-Mobile

A permanent (P) or temporary (T) place, or a stationary or mobile (M) facility, at or from which the organization performs activities forming part of its scope of accreditation, starting from sampling to final issuance of a report or certificate and / or quality system activities. A temporary (T) site is a site established under the responsibility of an accredited permanent site. All activities performed at a temporary site are the responsibility of the permanent site. An outdoors work is also considered to be a temporary site. Temporary site will be a site that involves work for special project and the activity will be defined in time (up to 2 years).
Type of Scopes: A- Fixed, C- Flexible scope in analytical tests : Type of matrix, analytes, experimental systems and/or analytical characteristics may be subject to changes, in accordance with the laboratory's approved and documented procedures. For details, please refer to the list of Accredited Tests, available from the laboratory upon request.



Item	Scope Type	Site	Measurand, Instrument, Gauge	Range [Including margins] (Does not include margins)	CMC Expressed as an Expanded Uncertainty (95%)	Reference Documents	Remarks	
Calibration – Electrical Quantities - DC and LF					כיוול – גדלים חשמליים - זרם ישר ותדר נמוך			
1	A	P	DC Voltage, Sources	מחוללי מתח בזרם ישר	[0 V to 0.1 V]	Manufacturer Instructions	Digital Multimeters: Datron 1271; Fluke 8508A	
2	A	P			(0.1 V to 0.2 V)			2 μV
3	A	P			(0.2 V to 1.0 V)			2.7 μV
4	A	P			(1.0 V to 2.0 V)			14 μV
5	A	P			(2.0 V to 10 V)			19 μV
6	A	P			(10 V to 20 V)			120 μV
7	A	P			(20 V to 100 V)			180 μV
8	A	P			(100 V to 200 V)			1.4 mV
9	A	P			(200 V to 1000 V)			2.3 mV
10	A	P	DC Voltage Reference Voltage Sources	מתח בזרם ישר, מקורות מתח ישר לייחוס	1.018 V	Manufacturer Instructions	Standard Cells Fluke 7001 Digital Multimeters: Datron 1271; Fluke 8508A	
11	A	P			10 V			2.2 μV
12	A	P	DC Voltage, Measuring Instruments	מכשירי מדידת מתח בזרם ישר	[0 V to 0.33 V]	Manufacturer Instructions	Calibrator Fluke 5520A	
13	A	P			(0.33 V to 3.3 V)			16 μV
14	A	P			(3.3 V to 33 V)			160 μV
15	A	P			(33 V to 330 V)			1.6 mV
16	A	P			(330 V to 1020 V)			21 mV
17	A	P	DC Current, Sources	זרם ישר, מחוללים	[0 mA to 0.2 mA]	Manufacturer Instructions	Digital Multimeters Datron 1271; Fluke 8508A	
18	A	P			(0.2 mA to 2.0 mA)			5.4 nA
19	A	P			(2.0 mA to 20 mA)			54 nA



Item	Scope Type	Site	Measurand, Instrument, Gauge		Range [Including margins] (Does not include margins)	CMC Expressed as an Expanded Uncertainty (95%)	Reference Documents	Remarks
Calibration – Electrical Quantities - DC and LF						כיוול – גדלים חשמליים - זרם ישר ותדר נמוך		
20	A	P			(20 mA to 200 mA)	17 μ A		
21	A	P			(200 mA to 1.0 A)	250 μ A		
22	A	P			(1.0 A to 2.0 A)	0.58 mA		
23	A	P			(2.0 A to 20 A)	13 mA		
24	A	P	DC Current, Measuring Instruments	זרם ישר, מכשירי מדידה	[0 mA to 0.32 mA]	0.7 μ A	Manufacturer Instructions	Calibrators Fluke 5520A;
25	A	P			(0.32 mA to 3.3 mA)	1.2 μ A		
26	A	P			(3.3 mA to 33 mA)	21 μ A		
27	A	P			(33 mA to 330 mA)	310 μ A		
28	A	P			(330 mA to 2.1 A)	1.2 mA		
29	A	P			(3.2 A to 11 A)	12 mA		
30	A	P	AC Voltage, Sources	מתח חילופין מחוללים	[40 Hz to 70 Hz]		Manufacturer Instructions	Digital Multimeters Datron 1271; Fluke 8508A
31	A	P			[0.00 V to 0.1 V]	0.053 mV		
32	A	P			(0.1 V to 0.2 V)	0.072 mV		
33	A	P			(0.2 V to 1.0 V)	0.2 mV		
34	A	P			(1.0 V to 2.0 V)	0.41 mV		
35	A	P			(2.0 V to 10 V)	3.8 mV		
36	A	P			(10 V to 20 V)	4.9 mV		
37	A	P			(20 V to 60 V)	6.0 mV		
					(60 V to 120 V)	11 mV		



Item	Scope Type	Site	Measurand, Instrument, Gauge	Range [Including margins] (Does not include margins)	CMC Expressed as an Expanded Uncertainty (95%)	Reference Documents	Remarks	
Calibration – Electrical Quantities - DC and LF					כיוול – גדלים חשמליים - זרם ישיר ותדר נמוך			
38	A	P		(120 V to 240 V)	22 mV			
39	A	P		(240 V to 480 V)	44 mV		Digital Multimeter Fluke 8508A	
40	A	P	AC Voltage. Measuring Instruments	מתח חילופין מכשירי מדידה	[40 Hz to 70 Hz]	Manufacturer Instructions	Calibrators Fluke 5520A;	
		[0.00 V to 0.01 V]			150 μ V			
41	A	P			(0.01 V to 0.033 V)			160 μ V
42	A	P			(0.033 V to 0.32 V)			600 μ V
43	A	P			(0.32 V to 3.2 V)			5.4 mV
44	A	P			(3.2 V to 33 V)			40 mV
45	A	P			(33 V to 330 V)			280 mV
46	A	P	(330 V to 1020 V)	1 V				
47	A	P	AC Voltage. Sources	מתח חילופין מחוללים	[70 Hz to 1 KHz]	Manufacturer Instructions	Digital Multimeters Datron 1271; Fluke 8508A	
		[0.00 V to 0.1 V]			0.053 mV			
48	A	P			(0.1 V to 0.2 V)			0.075 mV
49	A	P			(0.2 V to 1.0 V)			0.18 mV
50	A	P			(1.0 V to 2.0 V)			0.46 mV
51	A	P			(2.0 V to 10 V)			3.8 mV
52	A	P			(10 V to 20 V)			5.4 mV
53	A	P			(20 V to 100 V)			24 mV
54	A	P	(100 V to 200 V)	46 mV				



Item	Scope Type	Site	Measurand, Instrument, Gauge	Range [Including margins] (Does not include margins)	CMC Expressed as an Expanded Uncertainty (95%)	Reference Documents	Remarks
Calibration – Electrical Quantities - DC and LF					כיוול – גדלים חשמליים - זרם ישיר ותדר נמוך		
55	A	P		(200V to 1000V]	240 mV		
56	A	P	AC Current Sources	זרם חילופין מחוללים [40 Hz to 70 Hz]	0.073 μ A	Manufacturer Instructions	Digital Multimeters Datron 1271; Fluke 8508A
57	A	P		(100 μ A to 200 μ A]	0.17 μ A		
58	A	P		(200 μ A to 1.0 mA]	0.064 μ A		
59	A	P		(1.0 mA to 2.0 mA]	1.2 μ A		
60	A	P		(2.0 mA to 5.0 mA]	1.3 μ A		Three Phase Comparator Zera COM 303-1
61	A	P		(5.0 mA to 10 mA]	1.6 μ A		
62	A	P		(10 mA to 20 mA]	4.0 μ A		
63	A	P		(20 mA to 50 mA]	8.0 μ A		
64	A	P		(50 mA to 100 mA]	10 μ A		
65	A	P		(100mA to 200mA]	21 μ A		
66	A	P		(200mA to 500mA]	50 μ A		
67	A	P		(0.5 A to 1.0 A]	0.1 mA		
68	A	P		(1.0 A to 2.0 A]	0.21 mA		
69	A	P		(2.0 A to 5.0 A]]	0.5 mA		
70	A	P		(5.0 A to 10 A]	1.0 mA		
71	A	P		(10 A to 20 A]	4.0 mA		
72	A	P		(20 A to 50 A]	10 mA		



Item	Scope Type	Site	Measurand, Instrument, Gauge	Range [Including margins] (Does not include margins)	CMC Expressed as an Expanded Uncertainty (95%)	Reference Documents	Remarks
Calibration – Electrical Quantities - DC and LF					כיוול – גדלים חשמליים - זרם ישיר ותדר נמוך		
73	A	P		(50 A to 100 A)	18 mA		
74	A	P		(100 A to 160 A)	43 mA		
75	A	P		(160 A to 500 A)	157 mA		Three Phase Comparator Zera COM 303-1 Current Comparator Tettex 4761
76	A	P		(500 A to 1000 A)	197 mA		
77	A	P	AC Current Measuring Instruments	זרם חילופין מכשירי מדידה [40 Hz to 70 Hz]		Manufacturer Instructions	Calibrators Fluke 5520A;
				[0.0 μA to 33 μA]	0.92 μA		
78	A	P		(33 μA to 0.32 mA)	1.3 μA		
79	A	P		0.32mA to 3.2mA]	5.2 μA		
80	A	P		(3.2 mA to 33 mA)	56 μA		
81	A	P		(33 mA to 330 mA)	560 μA		
82	A	P		(0.33 A to 2.2 A)	4.2 mA		
83	A	P		(2.2 A to 11 A)	14 mA		
84	A	P		(11 A to 20 A)	70 mA		
85	A	P		(20 A to 32 A)	110 mA		
86	A	P		(32 A to 160 A)	560 mA		
87	A	P		(160 A to 200 A)	750 mA		
88	A	P		(200 A to 1000 A)	3.9 A		



Item	Scope Type	Site	Measurand, Instrument, Gauge	Range [Including margins] (Does not include margins)	CMC Expressed as an Expanded Uncertainty (95%)	Reference Documents	Remarks		
Calibration – Electrical Quantities - DC and LF					כיוול – גדלים חשמליים - זרם ישר ותדר נמוך				
89	A	P	AC Current Sources	זרם חילופין מחוללים [70 Hz to 1 KHz] [0.0μA to 0.1mA]	0.073 μA	Manufacturer Instructions	Digital Multimeters Datron 1271; Fluke 8508A		
90	A	P						(0.1 mA to 0.2mA)	0.2 μA
91	A	P						(0.2 mA to 1.0mA)	0.64 μA
92	A	P						(1.0 mA to 2.0mA)	1.4 μA
93	A	P						(2.0 mA to 10 mA)	6.4 μA
94	A	P						(10 mA to 20 mA)	14 μA
95	A	P						(20 mA to 100mA)	45 μA
96	A	P						(100mA to 200mA)	170 μA
97	A	P						(0.2 A to 1.0 A)	0.95 mA
98	A	P						(1.0 A to 2.0 A)	1.9 mA
99	A	P						(2.0 A to 20 A)	25 mA
100	A	P	DC resistance, Reference Resistance	התנגדות זרם ישר, התנגדויות לייחוס	13 μΩ	Manufacturer Instructions	Micro-Ohmmeter Cambridge 510A; Digital Multimeters Datron 1271; Fluke 8508A		
101	A	P						(0.02 Ω to 2.0 Ω)	77 μΩ
102	A	P						(2.0 Ω to 20 Ω)	0.37 mΩ
103	A	P						(20 Ω to 100 Ω)	1.9 mΩ
104	A	P						(100 Ω to 200 Ω)	3.5 mΩ
105	A	P						(200 Ω to 1.0 kΩ)	1.9 mΩ
106	A	P						(1.0 kΩ to 2.0 kΩ)	35 mΩ
107	A	P						(2.0 kΩ to 10 kΩ)	0.17 Ω



Item	Scope Type	Site	Measurand, Instrument, Gauge	Range [Including margins] (Does not include margins)	CMC Expressed as an Expanded Uncertainty (95%)	Reference Documents	Remarks	
Calibration – Electrical Quantities - DC and LF					כיוול – גדלים חשמליים - זרם ישיר ותדר נמוך			
108	A	P		(10 kΩ to 20 kΩ]	0.30 Ω			
109	A	P		(20 kΩ to 100 kΩ]	1.9 Ω			
110	A	P		(100 kΩ to 200 kΩ]	3.5 Ω			
111	A	P		(200 kΩ to 1.0 MΩ]	31 Ω			
112	A	P		(1.0 MΩ to 2.0 MΩ]	50 Ω			
113	A	P		(2.0 MΩ to 10 MΩ]	0.64 kΩ			
114	A	P		(10 MΩ to 20 MΩ]	0.93 kΩ			
115	A	P		(20 MΩ to 200 MΩ]	63 kΩ		Digital Multimeter Fluke 8508A	
116	A	P		(200 MΩ to 1 GΩ]	5.1 MΩ			
117	A	P	DC Resistance, Measuring Instruments	התנגדות זרם ישיר, מכשירי מדידה	(0.0 Ω to 11.0 Ω]	14 mΩ	Manufacturer Instructions	Calibrators Fluke 5520A;
118	A	P		(11.0 Ω to 33.0 Ω]	26 mΩ			
119	A	P		(33 Ω to 110 Ω]	40 mΩ			
120	A	P		(110 Ω to 330 Ω]	100 mΩ			
121	A	P		(330 Ω to 1.1 kΩ]	300 mΩ			
122	A	P		(1.1 kΩ to 3.3 kΩ]	1.0 Ω			
123	A	P		(3.3 kΩ to 11 kΩ]	3.0 Ω			
124	A	P		(11 kΩ to 33 kΩ]	7.0 Ω			
125	A	P		(33 kΩ to 110 kΩ]	28 Ω			
126	A	P		(110 kΩ to 330 kΩ]	85 Ω			



Item	Scope Type	Site	Measurand, Instrument, Gauge	Range [Including margins] (Does not include margins)	CMC Expressed as an Expanded Uncertainty (95%)	Reference Documents	Remarks
Calibration – Electrical Quantities - DC and LF					כיוול – גדלים חשמליים - זרם ישר ותדר נמוך		
127	A	P		(330 kΩ to 1.1 MΩ]	350 Ω		
128	A	P		(1.1 MΩ to 3.3 MΩ]	1.2 kΩ		
129	A	P		(3.3 MΩ to 11 MΩ]	12 kΩ		
130	A	P		(11 MΩ to 33 MΩ]	100 kΩ		
131	A	P		(33 MΩ to 110 MΩ]	920 kΩ		
132	A	P		(110 MΩ to 330 MΩ]	6.0 MΩ		
133	A	P	Phase angle, Wattmeter, Power source	זווית מופע, מד הספק, מקור הספק	[0° to 360°]	0.034°	Manufacturer Instructions Three Phase Comparator Zera COM 303-1
134	A	P1	Voltage ratio Voltage measuring transformer Primary: 6 kV to 40 kV Secondary: 110 V or 110/√3 V	יחס השנאה של שנאי מתח	1/300 to 1/120	2·10 ⁻⁴ V/V	IEC 60044-2 Reference V.T. MPS NVOD 30. Measuring Bridge Zera WM-303U 1' (minute) stands for 0.291 mrad
135	A	P1	Phase displacement Voltage measuring transformer Primary: 6 kV to 40 kV Secondary: 110 V or 110/√3 V	הזזת מופע של שנאי מתח	0' to 100'	2.0'	
136	A	P1	Current ratio Current measuring transformer Primary: 0.05 A to 1000 A Secondary: 5 A	יחס השנאה של שנאי זרם	1/200 to 1/1	1.1·10 ⁻⁴ A/A	IEC 60044-1 Current comparator TETTEX 4761. Measuring Bridge Zera WM-303I 1' (minute) stands for 0.291 mrad
137	A	P1	Phase displacement Current measuring transformer Primary: 0.05A to 1000A Secondary: 5A	הזזת מופע של שנאי זרם	0' to 100'	0.3'	



Item	Scope Type	Site	Measurand, Instrument, Gauge	Range [Including margins] (Does not include margins)	CMC Expressed as an Expanded Uncertainty (95%)	Reference Documents	Remarks
Calibration – Electrical Quantities - DC and LF					כיוול – גדלים חשמליים - זרם ישר ותדר נמוך		
138	A	P	Frequency, Sources	תדר, מחוללים [0.5 Hz to 10.0 MHz]	1.5 Hz / MHz	Manufacturer Instructions	Timer/ Counter Pendulum CNT 80
139	A	P	Frequency, Counters	תדר, מוני תדר [0.5 Hz to 10.0 MHz]	2.5 Hz / MHz	Manufacturer Instructions	Function Generator Philips PM 5193

Item	Scope Type	Site	Measurand Instrument, Gauge	Range [Including margins] (Does not include margins)	CMC Expressed as an Expanded Uncertainty (95%)	Reference Documents	Remarks
Calibration – Physical Quantities - Humidity					כיוול – גדלים פיזיקליים - לחות		
140	A	P	Relative Humidity, Hygrometers and humidity transducers, Ambient Humidity measurement	לחות יחסית, מדי לחות ומתמרים, מדידת לחות סביבה 23 °C [30 % rh to 95 % rh] 50 °C [10 % rh to 95 % rh]	6.0 % rh 7.0 % rh	IEC 60068 A Guide to the Measurement of humidity- NPL	Reference instrument TESTO 650



Item	Scope Type	Site	Measurand Instrument, Gauge	Range [Including margins] (Does not include margins)	CMC Expressed as an Expanded Uncertainty (95%)	Reference Documents	Remarks	
Calibration – Electrical Quantities - LF Single & Three Phase Electrical Power and Energy					כיוול – גדלים חשמליים - הספק ואנרגיה חשמלית חד ותלת מופעית בתדר נמוך			
141	A	P	Active Energy, in single phase 50 Hz network, אנרגיה אקטיבית ברשת חד מופעית 50 Hz	63.5 V	200 μ J/J	IEC 60736 IEC 60514 IEC 61358 IEC 60687 IEC 62052-11 IEC 62053-11 IEC 62053-21 IEC 62053-22 IEC 62053-23 IEC 60688 IEC 60145	Reference instrument COM 303 standard instrument Voltage measured phase to neutral	
				P. Factor 0.5				
				5 A				
142	A	P		P. Factor 1				150 μ J/J
				5 A				
143	A	P		10 A				150 μ J/J
144	A	P		230 V				
				P. Factor 0.5				
				2.5 A				200 μ J/J
145	A	P		5 A				200 μ J/J
146	A	P		10 A				200 μ J/J
147	A	P		20 A				220 μ J/J
148	A	P		50 A				220 μ J/J
149	A	P		100 A				220 μ J/J
150	A	P		P. Factor 1				
				0.05 A				170 μ J/J
151	A	P		0.1 A				150 μ J/J
152	A	P	0.2 A	150 μ J/J				
153	A	P	0.25 A	150 μ J/J				
154	A	P	0.5 A	150 μ J/J				
155	A	P	1 A	150 μ J/J				
156	A	P	2 A	150 μ J/J				
157	A	P	2.5 A	150 μ J/J				



Item	Scope Type	Site	Measurand Instrument, Gauge	Range [Including margins] (Does not include margins)	CMC Expressed as an Expanded Uncertainty (95%)	Reference Documents	Remarks
Calibration – Electrical Quantities - LF Single & Three Phase Electrical Power and Energy					כיוול – גדלים חשמליים - הספק ואנרגיה חשמלית חד ותלת מופעית בתדר נמוך		
158	A	P		5 A	150 μ J/J		
159	A	P		10 A	150 μ J/J		
160	A	P		20 A	170 μ J/J		
161	A	P		50 A	170 μ J/J		
162	A	P		100 A	170 μ J/J		
163	A	P	Active Energy, in three phase 50 Hz network, 3 wires, 2 elements	110 V P. Factor 0.5		IEC 60736 IEC 60514 IEC 61358 IEC 60687 IEC 62052-11 IEC 62053-11 IEC 62053-21 IEC 62053-22	Reference instrument COM 303 of Zera standard instrument. Voltage phase to phase
164	A	P		0.01 A	220 μ J/J		
165	A	P		0.02 A	220 μ J/J		
166	A	P		0.05 A	220 μ J/J		
167	A	P		0.1 A	200 μ J/J		
168	A	P		0.2 A	200 μ J/J		
169	A	P		0.5 A	200 μ J/J		
170	A	P		1 A	200 μ J/J		
171	A	P		2 A	200 μ J/J		
172	A	P		5 A	200 μ J/J		
173	A	P		P. Factor 1 0.01 A	170 μ J/J		
174	A	P		0.02 A	170 μ J/J		
				0.05 A	170 μ J/J		



Item	Scope Type	Site	Measurand Instrument, Gauge		Range [Including margins] (Does not include margins)	CMC Expressed as an Expanded Uncertainty (95%)	Reference Documents	Remarks
Calibration – Electrical Quantities - LF Single & Three Phase Electrical Power and Energy						כיול – גדלים חשמליים - הספק ואנרגיה חשמלית חד ותלת מופעית בתדר נמוך		
175	A	P			0.1 A	150 μ J/J		
176	A	P			0.2 A	150 μ J/J		
177	A	P			0.5 A	150 μ J/J		
178	A	P			1 A	150 μ J/J		
179	A	P			2 A	150 μ J/J		
180	A	P			5 A	150 μ J/J		
181	A	P			10 A	150 μ J/J		
182	A	P			20 A	170 μ J/J		
183	A	P	Active energy, in three phase 50 Hz network, 4 wires, 3 elements	אנרגיה אקטיבית ברשת תלת 50 Hz מופעית 4 מוליכים, 3 אלמנטים	63.5 V P. Factor 0.5			Reference instrument ILM 03 standard instrument Voltage measured phase to neutral
184	A	P			0.05 A	220 μ J/J		
185	A	P			0.1 A	200 μ J/J		
186	A	P			0.2 A	200 μ J/J		
187	A	P			0.5 A	200 μ J/J		
188	A	P			1 A	200 μ J/J		
189	A	P			2 A	200 μ J/J		
190	A	P			5 A	200 μ J/J		
191	A	P			10 A	200 μ J/J		
192	A	P			P. Factor 1 0.05 A 0.1 A	170 μ J/J 150 μ J/J		



Item	Scope Type	Site	Measurand Instrument, Gauge	Range [Including margins] (Does not include margins)	CMC Expressed as an Expanded Uncertainty (95%)	Reference Documents	Remarks
Calibration – Electrical Quantities - LF Single & Three Phase Electrical Power and Energy					כיוול – גדלים חשמליים - הספק ואנרגיה חשמלית חד ותלת מופעית בתדר נמוך		
193	A	P		0.2 A	150 μ J/J		
194	A	P		0.5 A	150 μ J/J		
195	A	P		1 A	150 μ J/J		
196	A	P		2 A	150 μ J/J		
197	A	P		5 A	150 μ J/J		
198	A	P		10 A	150 μ J/J		
199	A	P		230 V P. Factor 0.5			
200	A	P		0.01 A	220 μ J/J		
201	A	P		0.02 A	220 μ J/J		
202	A	P		0.05 A	220 μ J/J		
203	A	P		0.1 A	200 μ J/J		
204	A	P		0.2 A	200 μ J/J		
205	A	P		0.5 A	200 μ J/J		
206	A	P		1 A	200 μ J/J		
207	A	P		2 A	200 μ J/J		
208	A	P		5 A	200 μ J/J		
209	A	P		10 A	200 μ J/J		
210	A	P		20 A	220 μ J/J		
				50 A	220 μ J/J		



Item	Scope Type	Site	Measurand Instrument, Gauge	Range [Including margins] (Does not include margins)	CMC Expressed as an Expanded Uncertainty (95%)	Reference Documents	Remarks
Calibration – Electrical Quantities - LF Single & Three Phase Electrical Power and Energy					כיוול – גדלים חשמליים - הספק ואנרגיה חשמלית חד ותלת מופעית בתדר נמוך		
211	A	P		100 A	220 μ J/J		
212	A	P		P. Factor 1			
213	A	P		0.01 A	170 μ J/J		
214	A	P		0.02 A	170 μ J/J		
215	A	P		0.05 A	170 μ J/J		
216	A	P		0.1 A	150 μ J/J		
217	A	P		0.2 A	150 μ J/J		
218	A	P		0.5 A	150 μ J/J		
219	A	P		1 A	150 μ J/J		
220	A	P		2 A	150 μ J/J		
221	A	P		5 A	150 μ J/J		
222	A	P		10 A	150 μ J/J		
223	A	P		20 A	170 μ J/J		
224	A	P		50 A	170 μ J/J		
225	A	P	Reactive energy, in three phase	110 V			Reference instrument COM 303 of Zera standard instrument. Voltage phase to phase
226	A	P	50 Hz network, 3 wires, 2 elements	P. Factor 0.5	300 μ J/J		
227	A	P		0.05 A	280 μ J/J		
228	A	P		0.1 A	280 μ J/J		
	A	P		0.2 A	280 μ J/J		
	A	P		0.5 A	280 μ J/J		



Item	Scope Type	Site	Measurand Instrument, Gauge	Range [Including margins] (Does not include margins)	CMC Expressed as an Expanded Uncertainty (95%)	Reference Documents	Remarks
Calibration – Electrical Quantities - LF Single & Three Phase Electrical Power and Energy					כיוול – גדלים חשמליים - הספק ואנרגיה חשמלית חד ותלת מופעית בתדר נמוך		
229	A	P		1 A	280 μ J/J		
230	A	P		2 A	280 μ J/J		
231	A	P		5 A	280 μ J/J		
232	A	P		10 A	280 μ J/J		
233	A	P		P. Factor 1 0.01 A	240 μ J/J		
234	A	P		0.02 A	240 μ J/J		
235	A	P		0.05 A	240 μ J/J		
236	A	P		0.1 A	230 μ J/J		
237	A	P		0.2 A	230 μ J/J		
238	A	P		0.5 A	230 μ J/J		
239	A	P		1 A	230 μ J/J		
240	A	P		2 A	230 μ J/J		
241	A	P		5 A	230 μ J/J		
242	A	P		10 A	230 μ J/J		
243	A	P	Reactive energy, in three phase 50 Hz network, 4 wires, 3 elements	63.5 V P. Factor 0.5 0.05 A	300 μ J/J		Reference instrument COM 303 of Zera standard instrument. Voltage phase to neutral
244	A	P		0.1 A	280 μ J/J		
245	A	P		0.2 A	280 μ J/J		
246	A	P		0.5 A	280 μ J/J		



Item	Scope Type	Site	Measurand Instrument, Gauge	Range [Including margins] (Does not include margins)	CMC Expressed as an Expanded Uncertainty (95%)	Reference Documents	Remarks
Calibration – Electrical Quantities - LF Single & Three Phase Electrical Power and Energy					כיוול – גדלים חשמליים - הספק ואנרגיה חשמלית חד ותלת מופעית בתדר נמוך		
247	A	P		1 A	280 μ J/J		
248	A	P		2 A	280 μ J/J		
249	A	P		5 A	280 μ J/J		
250	A	P		10 A	280 μ J/J		
251	A	P		P. Factor 1 0.01 A	240 μ J/J		
252	A	P		0.02 A	240 μ J/J		
253	A	P		0.05 A	240 μ J/J		
254	A	P		0.1 A	230 μ J/J		
255	A	P		0.2 A	230 μ J/J		
256	A	P		0.5 A	230 μ J/J		
257	A	P		1 A	230 μ J/J		
258	A	P		2 A	230 μ J/J		
259	A	P		5 A	230 μ J/J		
260	A	P		10 A	230 μ J/J		
261	A	P	230 V	P. Factor 0.5 0.05 A	300 μ J/J		
262	A	P		0.1 A	280 μ J/J		
263	A	P		0.2 A	280 μ J/J		



Item	Scope Type	Site	Measurand Instrument, Gauge	Range [Including margins] (Does not include margins)	CMC Expressed as an Expanded Uncertainty (95%)	Reference Documents	Remarks
Calibration – Electrical Quantities - LF Single & Three Phase Electrical Power and Energy					כיוול – גדלים חשמליים - הספק ואנרגיה חשמלית חד ותלת מופעית בתדר נמוך		
264	A	P		0.5 A	280 μ J/J		
265	A	P		1 A	280 μ J/J		
266	A	P		2 A	280 μ J/J		
267	A	P		5 A	280 μ J/J		
268	A	P		10 A	280 μ J/J		
269	A	P		P. Factor 1			
				0.01 A	240 μ J/J		
270	A	P		0.02 A	240 μ J/J		
271	A	P		0.05 A	240 μ J/J		
272	A	P		0.1 A	230 μ J/J		
273	A	P		0.2 A	230 μ J/J		
274	A	P		0.5 A	230 μ J/J		
275	A	P		1 A	230 μ J/J		
276	A	P		2 A	230 μ J/J		
277	A	P		5 A	230 μ J/J		
278	A	P		10 A	230 μ J/J		



Item	Scope Type	Site	Measurand Instrument, Gauge	Range [Including margins] (Does not include margins)	CMC Expressed as an Expanded Uncertainty (95%)	Reference Documents	Remarks
Calibration – Physical Quantities - Temperature					כיוול – גדלים פיזיקליים - טמפרטורה		
279	A	P	Temperature, Temperature Meters and Transducers	טמפרטורה מדי טמפרטורה ומתמרים [0 °C to 60 °C]	0.5 °C	IEC 60068	Reference instrument TESTO 650