הרשות הלאומית להסמכת מעבדות
Israel Laboratory Accreditation Authority

ISO/IEC 17025:2017

מעבדות בדיקה

תעודת הסמכה מס' 0410

רפאל תעשיות מגופים בע"מ - מעבדת מדי מים

כתובת אתר ייחוס: רח' האילן 16, אור עקיבא, 30600

בתוקף מיום: 16.08.2026 **עד יום: 08.10.2028**

הארגון נבדק ונבחן על ידי הרשות הלאומית להסמכת מעבדות (להלן הרשות) ונמצא ראוי להסמכה בהתאם לנספח פירוט היקף ההסמכה המצורף לתעודה זו, המהווה חלק בלתי נפרד ממנה ומספרו זהה למספר התעודה.

הסמכה מצביעה על כשירות מקצועית ותפעול מערכת ניהול איכות בעלת הכרה בינלאומית.

הארגון המוסמך על ידי הרשות, עומד בתקנים/ בדרישות המפורטים מעלה. דרישות התקנים הם לכשירות מקצועית ולמערכות ניהול, שהינן הכרחיות למתן תוצאות אמינות. הסמכה זו ניתנה בהתאם לכללי ISO/IEC 17011:2017 לפיהם פועלת הרשות ובמסגרתם מקיימת פיקוח שוטף על הארגון לצורך בחינת תפקודו המתמשך בהתאם לדרישות ההסמכה.

ההסמכה תקפה כל עוד הארגון עונה לאמות המידה שנקבעו על ידי הרשות.

הרשות חתומה על הסכם הכרה רב צדדי (MLA) מול ארגון European Accreditation Cooperation (EA).

תעודה זו אינה מהווה אישור לפי סעיף 12 לחוק התקנים.

אתי פלר
מנכ"ל
הרשות הלאומית להסמכת מעבדות

תאריך הסמכה ראשון: 09.10.2018

הרשות הלאומית להסמכת מעבדות
Israel Laboratory Accreditation Authority

Testing Laboratories

ISO/IEC 17025:2017

Accreditation Certificate No. 0410

Raphael Valve Industries Ltd. - Water Meter Laboratory

Main site address: 16 HaIlan st., Or Akiva, 30600, Israel

Valid from: 16.08.2026 **Until: 08.10.2028**

The organization was assessed by the Israel Laboratory Accreditation Authority (ISRAC) and found to be worthy of accreditation to the detailed schedule attached.
The schedule is an integral part of this certificate and is numbered with the above certificate number.
Accreditation demonstrates technical competence and operation of an internationally recognized quality management system.
The organization accredited by ISRAC complies with the standards/requirements mentioned above, meets the technical competence requirements and management system requirements that are necessary for it to consistently deliver technically competent results. This accreditation is granted in accordance with the requirements of ISO/IEC 17011:2017, and entails periodic surveillance and reassessment by ISRAC to ensure that the organization continues to comply with the accreditation requirements.
The accreditation is valid provided that the organization continues to meet the criteria as laid down by ISRAC.
ISRAC is an EA-MLA (European Accreditation Cooperation Multi-Lateral Agreement) signatory.

This certificate does not constitute an approval in accordance with article 12 of the standard law.

Date of first accreditation: 09.10.2018

Etty Feller
General Manager
Israel Laboratory Accreditation Authority

***Name and Address*:**

Organization name	**Rafael Valves Industries Ltd. - Water Meter Laboratory**
Address	16 Hailan St., Or Akiva, 30600, Israel
Phone	04-6263555
Fax	04-6263552
E-mail	kguryev@talis-group.com

Site: P or T or M , P-Permanent, T-Temporary, M-Mobile
A permanent (P) or temporary (T) place, or a stationary or mobile (M) facility, at or from which the organization performs activities forming part of its scope of accreditation, starting from sampling to final issuance of a report or certificate and / or quality system activities. A temporary (T) site is a site established under the responsibility of an accredited permanent site. All activities performed at a temporary site are the responsibility of the permanent site. An outdoors work is also considered to be a temporary site. Temporary site will be a site that involves work for special project and the activity will be defined in time (up to 2 years).
Type of Scopes: A- Fixed, C- Flexible scope in analytical tests : Type of matrix, analytes, experimental systems and/or analytical characteristics may be subject to changes, in accordance with the laboratory's approved and documented procedures. For details, please refer to the list of Accredited Tests, available from the laboratory upon request.

Item	*Scope Type*	*Site*	*Materials / Products Tested*		*Types of Test / Properties Measured*		*Standard / Method*	*Opinion and Interpretation*	*Remarks*
Group of products: ***Industrial Products - Engineering Testing***									***משפחת מוצרים: מוצרי תעשייה - בדיקות הנדסיות***
Water Meters Testing									***בדיקת מדי מים***
1	A	P	Water Meter	מד מים	Determination of the intrinsic errors (of indication)	קביעת סטיות פנימיות (אינדיקציה)	ISO 4064-2 part 7.4 SI 40642 section 7.4	–	ISO= International Organization for Standardization SI=Standard of Israel 6 flow rate ranges Test bench 34 Flow 50 l/h to 570 m^3/h 2” to 8”
2	A	P	Water Meter	מד מים	Determination of the intrinsic errors (of indication)	קביעת סטיות פנימיות (אינדיקציה)	ISO 4064-2 part 7.4 SI 40642 section 7.4	-	Test bench 35 Flow 50 l/h to 60 m^3/h 1.5” to 2.5”
3	A	P	Water Meter	מד מים	Static Pressure Test	בדיקה לחץ סטטי	ISO 4064-2 part 7.3 SI 40642 section 7.4	–	
4	A	P	Water Meter	מד מים	Pressure loss test	בדיקה הפסד עומד	ISO 4064-2 part 7.9 SI 40642 section 7.4	–	