



ISO/IEC 17020: 2012, Type A

פעילות פיקוח

תעודת הסמכה מס' 563
ס.וי.- טסט בע"מ

כתובת אתר ייחוס: רח' טבנקין 6, בת ים, 595525

עד יום: 30.07.2024

בתוקף מיום: 31.07.2022

הארגון נבדק ונבחן על ידי הרשות הלאומית להסמכת מעבדות (להלן הרשות) ונמצא ראוי להסמכה בהתאם לנספח פירוט היקף ההסמכה המצורף לתעודה זו, המהווה חלק בלתי נפרד ממנה ומספרו זהה למספר התעודה. הסמכה מצביעה על כשירות מקצועית ותפעול מערכת ניהול איכות בעלת הכרה בינלאומית. הארגון המוסמך על ידי הרשות, עומד בתקנים/ בדרישות המפורטים מעלה. דרישות התקנים הם לכשירות מקצועית ולמערכות ניהול, שהינן הכרחיות למתן תוצאות אמינות. הסמכה זו ניתנה בהתאם לכללי ISO/IEC 17011:2017 לפיהם פועלת הרשות ובמסגרתם מקיימת פיקוח שוטף על הארגון לצורך בחינת תפקודו המתמשך בהתאם לדרישות ההסמכה. ההסמכה תקפה כל עוד הארגון עונה לאמות המידה שנקבעו על ידי הרשות. הרשות חתומה על הסכם הכרה רב צדדי (MLA) מול ארגון European Accreditation Cooperation (EA).

תעודה זו אינה מהווה אישור לפי סעיף 12 לחוק התקנים.

תאריך הסמכה ראשון: 31.07.2022

אתי פלר
מנכ"ל
הרשות הלאומית להסמכת מעבדות



Inspection Activity

ISO/IEC 17020: 2012, Type A

Accreditation Certificate No. 563
S.V.-TEST LTD

Main site address: 6 Tabenkin St., Bat Yam, 595525, Israel

Valid from: 31.07.2022

Until: 30.07.2024

The organization was assessed by the Israel Laboratory Accreditation Authority (ISRAC) and found to be worthy of accreditation to the detailed schedule attached.

The schedule is an integral part of this certificate and is numbered with the above certificate number.

Accreditation demonstrates technical competence and operation of an internationally recognized quality management system.

The organization accredited by ISRAC complies with the standards/requirements mentioned above, meets the technical competence requirements and management system requirements that are necessary for it to consistently deliver technically competent results. This accreditation is granted in accordance with the requirements of ISO/IEC 17011:2017, and entails periodic surveillance and reassessment by ISRAC to ensure that the organization continues to comply with the accreditation requirements.

The accreditation is valid provided that the organization continues to meet the criteria as laid down by ISRAC. ISRAC is an EA-MLA (European Accreditation Cooperation Multi-Lateral Agreement) signatory.

This certificate does not constitute an approval in accordance with article 12 of the standard law.

Etty Feller
General Manager
Israel Laboratory Accreditation Authority

Date of first accreditation: 31.07.2022

Date of signature 17/01/2023

Page No. 2 of: 4



Name and Address:

Organization name **S.V.-TEST LTD**
Address 6 Tabenkin St., Bat Yam, 595525, Israel
Phone 053-7888898
Fax
E-mail (contact person) sergey.svtest@gmail.com

Site: P or T or M , P-Permanent, T-Temporary, M-Mobile

A permanent (P) or temporary (T) place, or a stationary or mobile (M) facility, at or from which the organization performs activities forming part of its scope of accreditation, starting from sampling to final issuance of a report or certificate and / or quality system activities. A temporary (T) site is a site established under the responsibility of an accredited permanent site. All activities performed at a temporary site are the responsibility of the permanent site. An outdoors work is also considered to be a temporary site. Temporary site will be a site that involves work for special project and the activity will be defined in time (up to 2 years).

Type of Scopes: A- Fixed, C- Flexible scope in analytical tests : Type of matrix, analytes, experimental systems and/or analytical characteristics may be subject to changes, in accordance with the laboratory's approved and documented procedures. For details, please refer to the list of Accredited Tests, available from the laboratory upon request.



Item	Site	Field	Requirements	Standard / Method	Remarks
Group of Products: Fire & Rescue – Fire Detection Systems, Alternative Extinguishing Systems				משפחת מוצרים: כבאות והצלה - מערכות גילוי אש וכיבוי אלטרנטיבי	
Proper Installation				בדיקת תקינות התקנה	
1	T	Fire detection systems	מערכות גילוי אש Installation instructions and general requirements	הוראות התקנה ודרישות כלליות	SI 1220 part 3
2	T	Fire detection systems	מערכות גילוי אש Operating Regulations Fire Safety Systems - Integration	משטר הפעלות מערכות בטיחות אש – אינטגרציה	הוראת נציב כבאות והצלה מסי 536
3	T	Clean agent automatic fire extinguishing systems	מערכות כיבוי אש אוטומטיות בגז כיבוי	התקנה	SI 1597
4	T	Aerosol fire extinguishing systems	מערכות לכיבוי אש בארוסול	התקנה	SI 5210

Item	Site	Field	Requirements	Standard / Method	Remarks
Group of Products: Fire & Rescue – Fire Detection Systems, Alternative Extinguishing Systems				משפחת מוצרים: כבאות והצלה - מערכות גילוי אש וכיבוי אלטרנטיבי	
System Planning				בדיקת תכנון	
5	T	Aerosol fire extinguishing systems	מערכות לכיבוי אש בארוסול Design	תכן	SI 5210
6	T	Clean agent automatic fire extinguishing systems	מערכות כיבוי אש אוטומטיות בגז כיבוי Design	תכן	SI 1597